

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 01/26/2018

DATE & TIME IN: 01/26/2018 08:30

OVEN TEMP: 103 °C

Run Number: LB92886

DATE & TIME OUT: 01/26/2018 13:20

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB92886BL	LB92886BL	1.8852	1.8852	100	1.8852	1.8852	1.8852	0	0
2	LB92886BS	LB92886BS	1.88	1.88	100	1.9346	1.9346	1.9346	0.0546	546
3	J1329-01	SPDES-1	1.8729	1.8729	800	1.8745	1.8745	1.8745	0.0016	2

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS012618

WorkList ID : 107904

Date : 1/26/2018 7:44:04 AM

61892886

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/26/2018	Water	J1329-01	TSS	Cool 4 deg C	SKAN06	N22	SPDES-1	01/25/2018	SM2540 D

Date/Time 01/26/18 8:00
 Received by: AS
 Relinquished by: CQ

Date/Time 01/26/18 8:45
 Received by: CQ
 Relinquished by: AS