


7470A

LB 92891
INSTRUMENT ID: CY1

Linear

 μ Abs.:

39629



A= 0.0000e+000

B= 2.5326e-004

C= -1.9123e-002

Rho= 0.9999650

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.039	0.039	231	0.000	231				
0.2	0.200	0.212	0.012	913	0.0 %	913				
2.5	2.500	2.447	-0.053	9737	0.0 %	9737				
5.0	5.000	4.972	-0.028	19709	0.0 %	19709				
7.5	7.500	7.512	0.012	29736	0.0 %	29736				
10.0	10.000	10.017	0.017	39629	0.0 %	39629				

LB92891 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Jan 2018 08:59:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	231	-	0.0000	7470A	PPB	26 Jan 2018 09:30:45	S	Std
0.2 - 1	50.2	913	-	0.2000	7470A	PPB	26 Jan 2018 09:32:53	S	Std
2.5 - 1	52.5	9737	-	2.5000	7470A	PPB	26 Jan 2018 09:35:01	S	Std
5.0 - 1	55	19709	-	5.0000	7470A	PPB	26 Jan 2018 09:37:10	S	Std
7.5 - 1	57.5	29736	-	7.5000	7470A	PPB	26 Jan 2018 09:39:17	S	Std
10.0 - 1	510	39629	-	10.0000	7470A	PPB	26 Jan 2018 09:41:25	S	Std
ICV66 - 1	ICV66	14680	3.6988	-	7470A	PPB	26 Jan 2018 09:44:06	U	SMPL
ICB66 - 1	ICB66	-135	-0.0533	-	7470A	PPB	26 Jan 2018 09:46:12	U	SMPL
CCV74 - 1	CCV74	20088	5.0684	-	7470A	PPB	26 Jan 2018 09:48:19	U	SMPL
CCB74 - 1	CCB74	-151	-0.0574	-	7470A	PPB	26 Jan 2018 09:50:26	U	SMPL
CRA - 1	CRA	941	0.2192	-	7470A	PPB	26 Jan 2018 09:52:33	U	SMPL
HighStd - 1	HighStd	39868	10.0779	-	7470A	PPB	26 Jan 2018 09:54:40	U	SMPL
ChkStd - 1	ChkStd	26268	6.6335	-	7470A	PPB	26 Jan 2018 09:56:47	U	SMPL
PB106025BL - 1	PBW	-171	-0.0624	-	7470A	PPB	26 Jan 2018 09:58:56	U	SMPL
PB106025BS - 1	LCSW	14751	3.7167	-	7470A	PPB	26 Jan 2018 10:01:04	U	SMPL
PB106025TB - 1	PB105978TB	262	0.0472	-	7470A	PPB	26 Jan 2018 10:03:12	U	SMPL
J1289-02 - 1	SB-08-COMP	326	0.0634	-	7470A	PPB	26 Jan 2018 10:05:20	U	SMPL
J1289-02DUP - 1	SB-08-COMP DUP	354	0.0705	-	7470A	PPB	26 Jan 2018 10:07:28	U	SMPL
J1289-02MS - 1	SB-08-COMPMS	16075	4.0521	-	7470A	PPB	26 Jan 2018 10:09:36	U	SMPL
J1289-02MSD - 1	SB-08-COMPMSD	16204	4.0847	-	7470A	PPB	26 Jan 2018 10:11:43	U	SMPL
CCV75 - 1	CCV75	19853	5.0089	-	7470A	PPB	26 Jan 2018 10:13:50	U	SMPL
CCB75 - 1	CCB75	-149	-0.0569	-	7470A	PPB	26 Jan 2018 10:15:57	U	SMPL
J1289-04 - 1	SB-09-COMP	301	0.0571	-	7470A	PPB	26 Jan 2018 10:18:04	U	SMPL
J1289-06 - 1	SB-10-COMP	279	0.0515	-	7470A	PPB	26 Jan 2018 10:20:11	U	SMPL
J1289-08 - 1	SB-11-COMP	316	0.0609	-	7470A	PPB	26 Jan 2018 10:22:17	U	SMPL
J1289-10 - 1	SB-60-COMP	287	0.0536	-	7470A	PPB	26 Jan 2018 10:24:24	U	SMPL
J1289-12 - 1	SB-56-COMP	274	0.0503	-	7470A	PPB	26 Jan 2018 10:26:32	U	SMPL
J1289-14 - 1	SB-53-COMP	316	0.0609	-	7470A	PPB	26 Jan 2018 10:28:40	U	SMPL
J1289-16 - 1	SB-52-COMP	303	0.0576	-	7470A	PPB	26 Jan 2018 10:30:48	U	SMPL
J1289-18 - 1	SB-54-COMP	259	0.0465	-	7470A	PPB	26 Jan 2018 10:32:56	U	SMPL
J1289-20 - 1	SB-55-COMP	274	0.0503	-	7470A	PPB	26 Jan 2018 10:35:05	U	SMPL
J1290-02 - 1	SB-51-COMP	318	0.0614	-	7470A	PPB	26 Jan 2018 10:37:12	U	SMPL
CCV76 - 1	CCV76	19791	4.9932	-	7470A	PPB	26 Jan 2018 10:39:19	U	SMPL
CCB76 - 1	CCB76	111	0.0090	-	7470A	PPB	26 Jan 2018 10:44:02	U	SMPL
J1290-04 - 1	SB-50-COMP	303	0.0576	-	7470A	PPB	26 Jan 2018 10:46:08	U	SMPL
J1290-06 - 1	SB-48-COMP	7	-0.0173	-	7470A	PPB	26 Jan 2018 10:48:15	U	SMPL
J1290-08 - 1	SB-58-COMP	14	-0.0156	-	7470A	PPB	26 Jan 2018 10:50:22	U	SMPL
J1290-10 - 1	SB-59-COMP	6	-0.0176	-	7470A	PPB	26 Jan 2018 10:52:29	U	SMPL
J1290-17 - 1	SB-49-COMP	-31	-0.0270	-	7470A	PPB	26 Jan 2018 10:54:37	U	SMPL
J1290-19 - 1	SB-46-COMP	0	-0.0191	-	7470A	PPB	26 Jan 2018 10:56:44	U	SMPL
J1290-21 - 1	SB-45-COMP	35	-0.0103	-	7470A	PPB	26 Jan 2018 10:58:53	U	SMPL
J1290-23 - 1	SB-43-COMP	56	-0.0049	-	7470A	PPB	26 Jan 2018 11:01:01	U	SMPL
PB106026BL - 1	PBW	-5	-0.0204	-	7470A	PPB	26 Jan 2018 11:03:10	U	SMPL
PB106026BS - 1	LCSW	15123	3.8109	-	7470A	PPB	26 Jan 2018 11:05:17	U	SMPL
CCV77 - 1	CCV77	19668	4.9620	-	7470A	PPB	26 Jan 2018 11:07:24	U	SMPL
CCB77 - 1	CCB77	-64	-0.0353	-	7470A	PPB	26 Jan 2018 11:09:30	U	SMPL
PB106026TB - 1	PB105979TB	51	-0.0062	-	7470A	PPB	26 Jan 2018 11:11:38	U	SMPL
J1020-02 - 1	PL-01-012418-A	19	-0.0143	-	7470A	PPB	26 Jan 2018 11:13:45	U	SMPL
J1020-02DUP - 1	PL-01-012418-ADUP	72	-0.0009	-	7470A	PPB	26 Jan 2018 11:15:53	U	SMPL
J1020-02MS - 1	PL-01-012418-AMS	16147	4.0703	-	7470A	PPB	26 Jan 2018 11:18:00	U	SMPL
J1020-02MSD - 1	PL-01-012418-AMSD	15813	3.9857	-	7470A	PPB	26 Jan 2018 11:20:07	U	SMPL
J1020-04 - 1	PL-01-012418-B	-113	-0.0477	-	7470A	PPB	26 Jan 2018 11:22:14	U	SMPL
J1020-06 - 1	PL-01-012418-C	-22	-0.0247	-	7470A	PPB	26 Jan 2018 11:24:22	U	SMPL
J1020-08 - 1	PL-01-012418-D	28	-0.0120	-	7470A	PPB	26 Jan 2018 11:26:30	U	SMPL
J1020-10 - 1	PL-02-012418-A	52	-0.0060	-	7470A	PPB	26 Jan 2018 11:28:38	U	SMPL
J1020-12 - 1	PL-02-012418-B	245	0.0429	-	7470A	PPB	26 Jan 2018 11:30:45	U	SMPL
CCV78 - 1	CCV78	19061	4.8083	-	7470A	PPB	26 Jan 2018 11:32:52	U	SMPL
CCB78 - 1	CCB78	-197	-0.0690	-	7470A	PPB	26 Jan 2018 11:34:59	U	SMPL
J1020-14 - 1	PL-02-012418-C	267	0.0485	-	7470A	PPB	26 Jan 2018 11:37:06	U	SMPL
J1020-16 - 1	PL-02-012418-D	262	0.0472	-	7470A	PPB	26 Jan 2018 11:39:14	U	SMPL
I1300-05 - 1	TP-1-5	309	0.0591	-	7470A	PPB	26 Jan 2018 11:41:20	U	SMPL
I1300-07 - 1	MANHOLE	306	0.0584	-	7470A	PPB	26 Jan 2018 11:43:28	U	SMPL
I1306-04 - 1	WC1	245	0.0429	-	7470A	PPB	26 Jan 2018 11:45:35	U	SMPL
I1306-08 - 1	WC2	72	-0.0009	-	7470A	PPB	26 Jan 2018 11:47:43	U	SMPL
I1306-12 - 1	WC3	71	-0.0011	-	7470A	PPB	26 Jan 2018 11:49:51	U	SMPL
I1306-16 - 1	WC4	66	-0.0024	-	7470A	PPB	26 Jan 2018 11:51:59	U	SMPL
I1306-20 - 1	WC5	37	-0.0098	-	7470A	PPB	26 Jan 2018 11:54:08	U	SMPL
I1307-08 - 1	MANHOLE-202	33	-0.0108	-	7470A	PPB	26 Jan 2018 11:56:15	U	SMPL
CCV79 - 1	CCV79	18205	4.5915	-	7470A	PPB	26 Jan 2018 11:58:23	U	SMPL
CCB79 - 1	CCB79	44	-0.0080	-	7470A	PPB	26 Jan 2018 12:00:30	U	SMPL
I1289-02LX5 - 1		-27	-0.0260	-	7470A	PPB	26 Jan 2018 12:02:38	U	SMPL
I1289-02A - 1		15848	3.9946	-	7470A	PPB	26 Jan 2018 12:04:46	U	SMPL
I1020-02LX5 - 1		-185	-0.0660	-	7470A	PPB	26 Jan 2018 12:06:53	U	SMPL
I1020-02A - 1		15168	3.8223	-	7470A	PPB	26 Jan 2018 12:09:00	U	SMPL
CCV80 - 1	CCV80	20152	5.0846	-	7470A	PPB	26 Jan 2018 12:13:47	U	SMPL
CCB80 - 1	CCB80	32	-0.0110	-	7470A	PPB	26 Jan 2018 12:15:54	U	SMPL