

L392893


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

41163



A= 0.0000e+000

B= 2.4249e-004

C= 5.5524e-003

Rho= 0.9999781

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.029	-0.029	-141	0.000	-141				
0.2	0.200	0.212	0.012	851	0.0 %	851	0			
2.5	2.500	2.496	-0.004	10271	0.0 %	10271				
5.0	5.000	5.047	0.047	20792	0.0 %	20792				
7.5	7.500	7.486	-0.014	30849	0.0 %	30849				
10.0	10.000	9.987	-0.013	41163	0.0 %	41163				

LB92893 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Jan 2018 12:36:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-141	-	0.0000	7471B	PPB	26 Jan 2018 12:41:30	S	Std
0.2 - 1	50.2	851	-	0.2000	7471B	PPB	26 Jan 2018 12:43:48	S	Std
2.5 - 1	52.5	10271	-	2.5000	7471B	PPB	26 Jan 2018 12:46:06	S	Std
5.0 - 1	55	20792	-	5.0000	7471B	PPB	26 Jan 2018 12:48:23	S	Std
7.5 - 1	57.5	30849	-	7.5000	7471B	PPB	26 Jan 2018 12:50:40	S	Std
10.0 - 1	510	41163	-	10.0000	7471B	PPB	26 Jan 2018 12:52:57	S	Std
ICV67 - 1	ICV67	16797	4.0786	-	7471B	PPB	26 Jan 2018 12:55:57	U	SMPL
ICB67 - 1	ICB67	-160	-0.0332	-	7471B	PPB	26 Jan 2018 12:58:13	U	SMPL
CCV81 - 1	CCV81	20433	4.9603	-	7471B	PPB	26 Jan 2018 13:02:59	U	SMPL
CCB81 - 1	CCB81	-128	-0.0255	-	7471B	PPB	26 Jan 2018 13:05:14	U	SMPL
CRA - 1	CRA	863	0.2148	-	7471B	PPB	26 Jan 2018 13:07:31	U	SMPL
HighStd - 1	HighStd	41416	10.0485	-	7471B	PPB	26 Jan 2018 13:09:46	U	SMPL
ChkStd - 1	ChkStd	27906	6.7724	-	7471B	PPB	26 Jan 2018 13:12:03	U	SMPL
PB105926BL - 1	PBS	-317	-0.0713	-	7471B	PPB	26 Jan 2018 13:14:20	U	SMPL
PB105926BS - 1	LCSS	15669	3.8051	-	7471B	PPB	26 Jan 2018 13:16:38	U	SMPL
J1253-01 - 1	WC-01(0-5)	14219	3.4535	-	7471B	PPB	26 Jan 2018 13:18:55	U	SMPL
J1253-03 - 1	WC-02(0-5)	11601	2.8187	-	7471B	PPB	26 Jan 2018 13:21:13	U	SMPL
J1253-05 - 1	WC-03(0-5)	5896	1.4353	-	7471B	PPB	26 Jan 2018 13:23:30	U	SMPL
J1253-07 - 1	WC-04(0-5)	12139	2.9491	-	7471B	PPB	26 Jan 2018 13:25:48	U	SMPL
J1253-09 - 1	WC-05(0-5)	5496	1.3383	-	7471B	PPB	26 Jan 2018 13:28:04	U	SMPL
CCV82 - 1	CCV82	20501	4.9768	-	7471B	PPB	26 Jan 2018 13:33:07	U	SMPL
CCB82 - 1	CCB82	-176	-0.0371	-	7471B	PPB	26 Jan 2018 13:35:23	U	SMPL
J1253-11 - 1	WC-06(5-16)	10864	2.6399	-	7471B	PPB	26 Jan 2018 13:37:40	U	SMPL
J1253-13 - 1	WC-07(5-15)	23090	5.6046	-	7471B	PPB	26 Jan 2018 13:39:56	U	SMPL
J1253-15 - 1	WC-08(5-15)	8237	2.0029	-	7471B	PPB	26 Jan 2018 13:42:12	U	SMPL
J1253-17 - 1	WC-9(5-15)	6573	1.5994	-	7471B	PPB	26 Jan 2018 13:44:28	U	SMPL
J1253-19 - 1	WC-10(0-5)	161389	39.1406	-	7471B	PPB	26 Jan 2018 13:46:45	U	SMPL
J1257-01 - 1	WC-11(0-5)	91726	22.2481	-	7471B	PPB	26 Jan 2018 14:04:19	U	SMPL
J1257-01DUP - 1	WC-11(0-5)DUP	89957	21.8191	-	7471B	PPB	26 Jan 2018 14:06:35	U	SMPL
J1257-01MS - 1	WC-11(0-5)MS	104527	25.3522	-	7471B	PPB	26 Jan 2018 14:08:55	U	SMPL
J1257-01MSD - 1	WC-11(0-5)MSD	107771	26.1388	-	7471B	PPB	26 Jan 2018 14:14:15	U	SMPL
J1257-03 - 1	WC-12(0-5)	57101	13.8519	-	7471B	PPB	26 Jan 2018 14:16:34	U	SMPL
CCV83 - 1	CCV83	20662	5.0159	-	7471B	PPB	26 Jan 2018 14:18:52	U	SMPL
CCB83 - 1	CCB83	-173	-0.0364	-	7471B	PPB	26 Jan 2018 14:21:10	U	SMPL
J1257-05 - 1	WC-13(0-5)	281400	68.2419	-	7471B	PPB	26 Jan 2018 14:23:28	U	SMPL
J1257-07 - 1	WC-14(0-5)	21630	5.2506	-	7471B	PPB	26 Jan 2018 14:25:44	U	SMPL
J1257-09 - 1	WC-15(0-5)	51696	12.5412	-	7471B	PPB	26 Jan 2018 14:28:22	U	SMPL
J1257-11 - 1	WC-16(0-5)	31698	7.6920	-	7471B	PPB	26 Jan 2018 14:30:39	U	SMPL
J1257-13 - 1	WC-17(0-5)	11787	2.8638	-	7471B	PPB	26 Jan 2018 14:32:59	U	SMPL
J1257-15 - 1	WC-19(5-15)	66047	16.0212	-	7471B	PPB	26 Jan 2018 14:35:17	U	SMPL
J1257-17 - 1	WC-20(5-15)	1549	0.3812	-	7471B	PPB	26 Jan 2018 14:37:34	U	SMPL
J1279-01 - 1	PB-1(25-32)	424	0.1084	-	7471B	PPB	26 Jan 2018 14:39:54	U	SMPL
PB106035BL - 1	PBS	-134	-0.0269	-	7471B	PPB	26 Jan 2018 14:42:11	U	SMPL
PB106035BS - 1	LCSS	15452	3.7525	-	7471B	PPB	26 Jan 2018 14:44:28	U	SMPL
CCV84 - 1	CCV84	20953	5.0864	-	7471B	PPB	26 Jan 2018 14:46:44	U	SMPL
CCB84 - 1	CCB84	-121	-0.0238	-	7471B	PPB	26 Jan 2018 14:49:00	U	SMPL
I1009-01 - 1	AU-05-012518-A	1255	0.3099	-	7471B	PPB	26 Jan 2018 14:51:17	U	SMPL
I1009-03 - 1	AU-05-012518-B	1285	0.3172	-	7471B	PPB	26 Jan 2018 14:53:33	U	SMPL
I1009-05 - 1	AU-05-012518-C	716	0.1792	-	7471B	PPB	26 Jan 2018 14:55:50	U	SMPL
I1009-07 - 1	AU-05-012518-D	966	0.2398	-	7471B	PPB	26 Jan 2018 14:58:06	U	SMPL
I1009-09 - 1	AU-06-012518-A	1126	0.2786	-	7471B	PPB	26 Jan 2018 15:00:23	U	SMPL
I1009-11 - 1	AU-06-012518-B	1093	0.2706	-	7471B	PPB	26 Jan 2018 15:02:40	U	SMPL
I1009-13 - 1	AU-06-012518-C	1117	0.2764	-	7471B	PPB	26 Jan 2018 15:04:57	U	SMPL
I1010-08 - 1	BU-01-012518	825	0.2056	-	7471B	PPB	26 Jan 2018 15:07:14	U	SMPL
I1018-10 - 1	OR-03-012518	584	0.1472	-	7471B	PPB	26 Jan 2018 15:09:31	U	SMPL
I1023-05 - 1	TR-05-012518	1047	0.2594	-	7471B	PPB	26 Jan 2018 15:11:48	U	SMPL
CCV85 - 1	CCV85	20839	5.0588	-	7471B	PPB	26 Jan 2018 15:14:04	U	SMPL
CCB85 - 1	CCB85	-130	-0.0260	-	7471B	PPB	26 Jan 2018 15:16:20	U	SMPL
I1307-01 - 1	MANHOLE-201	798	0.1991	-	7471B	PPB	26 Jan 2018 15:18:37	U	SMPL
I1307-05 - 1	MANHOLE-202	1023	0.2536	-	7471B	PPB	26 Jan 2018 15:20:53	U	SMPL
I1320-01 - 1	T1WC-1	427	0.1091	-	7471B	PPB	26 Jan 2018 15:23:09	U	SMPL
I1322-01 - 1	M3-LEFT-RIGHT	-141	-0.0286	-	7471B	PPB	26 Jan 2018 15:25:26	U	SMPL
I1325-01 - 1	RO-4632	1049	0.2599	-	7471B	PPB	26 Jan 2018 15:27:42	U	SMPL
I1325-01DUP - 1	RO-4632DUP	1080	0.2674	-	7471B	PPB	26 Jan 2018 15:32:45	U	SMPL
I1325-01MS - 1	RO-4632MS	17184	4.1725	-	7471B	PPB	26 Jan 2018 15:35:01	U	SMPL
I1325-01MSD - 1	RO-4632MSD	17171	4.1693	-	7471B	PPB	26 Jan 2018 15:37:19	U	SMPL
I1325-04 - 1	RO-3025	380	0.0977	-	7471B	PPB	26 Jan 2018 15:39:37	U	SMPL
I1325-07 - 1	RO-2334	1016	0.2519	-	7471B	PPB	26 Jan 2018 15:41:53	U	SMPL
CCV86 - 1	CCV86	20996	5.0968	-	7471B	PPB	26 Jan 2018 15:44:09	U	SMPL
CCB86 - 1	CCB86	-125	-0.0248	-	7471B	PPB	26 Jan 2018 15:46:26	U	SMPL
I1325-10 - 1	RO-2342	218	0.0584	-	7471B	PPB	26 Jan 2018 15:48:43	U	SMPL
I1326-01 - 1	RO200030	1005	0.2493	-	7471B	PPB	26 Jan 2018 15:51:00	U	SMPL
I1328-01 - 1	RO-1-2	1134	0.2805	-	7471B	PPB	26 Jan 2018 15:53:17	U	SMPL
I1257-01LX5 - 1		17946	4.3573	-	7471B	PPB	26 Jan 2018 15:58:23	U	SMPL
I1257-01A - 1		108994	26.4354	-	7471B	PPB	26 Jan 2018 16:03:18	U	SMPL
I1325-01LX5 - 1		242	0.0642	-	7471B	PPB	26 Jan 2018 16:11:07	U	SMPL
I1325-01A - 1		17971	4.3633	-	7471B	PPB	26 Jan 2018 16:13:24	U	SMPL
I1253-19DLX10 - 1		15774	3.8306	-	7471B	PPB	26 Jan 2018 16:15:41	U	SMPL
I1257-03DLX5 - 1		11097	2.6964	-	7471B	PPB	26 Jan 2018 16:17:59	U	SMPL
I1257-05DLX10 - 1		29530	7.1662	-	7471B	PPB	26 Jan 2018 16:20:16	U	SMPL
CV87 - 1	CCV87	20324	4.9339	-	7471B	PPB	26 Jan 2018 16:22:33	U	SMPL

LB92893 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Jan 2018 12:36:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB87 - 1	CCB87	-205	-0.0442	-7471B	PPB		26 Jan 2018 16:24:51	U	SMPL
J1257-01DLX5 - 1		17965	4.3619	-7471B	PPB		26 Jan 2018 16:27:08	U	SMPL
J1257-01DUPDLX5 - 1		17149	4.1640	-7471B	PPB		26 Jan 2018 16:35:55	U	SMPL
J1257-01MSDLX5 - 1		21623	5.2489	-7471B	PPB		26 Jan 2018 16:44:24	U	SMPL
J1257-01MSDDLX5 - 1		20583	4.9967	-7471B	PPB		26 Jan 2018 16:46:40	U	SMPL
J1257-01LDLX25 - 1		3267	0.7978	-7471B	PPB		26 Jan 2018 16:48:57	U	SMPL
J1257-01ADLX5 - 1		21780	5.2870	-7471B	PPB		26 Jan 2018 16:56:47	U	SMPL
J1257-09DLX5 - 1		9153	2.2251	-7471B	PPB		26 Jan 2018 16:59:03	U	SMPL
J1257-15DLX5 - 1		11835	2.8754	-7471B	PPB		26 Jan 2018 17:01:21	U	SMPL
CCV88 - 1	CCV88	20786	5.0459	-7471B	PPB		26 Jan 2018 17:03:38	U	SMPL
CCB88 - 1	CCB88	-299	-0.0670	-7471B	PPB		26 Jan 2018 17:05:55	U	SMPL