

Analytical Summary Report

Analysis Method: 9038

ANALYST: AMANDEEP

Parameter: Sulfate

SUPERVISOR REVIEW BY: apatel

Run Number: LB92903

Reagent/Standard	Lot/Log #
sulfate ICV 20ppm	WP61288
Sulfate LCS std, 20ppm	WP61289
sulfate 0.0ppm	WP61279
sulfate 5ppm	WP61280
sulfate 10ppm	WP61281
sulfate 20ppm	WP61283
sulfate 30ppm	WP61285
sulfate 40ppm	WP61286
sulfate CCV 20ppm	WP61287
Sulfate std, 15ppm	WP61282
Sulfate std, 35ppm	WP61292
Sulfate std, 25ppm	WP61284
conditioning reagent	WP59618
BARIUM CHLORIDE DIHYDRATE ACS 100G	W2210

Intercept: -1.7996

Slope: 3.1849

Regression: 0.999528

Seq	Lab ID	True Val (mg/l)	DF	MATRIX	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	Water	0.201	0.396	0.195	02/19/2018	09:58
2	CAL2	5	1	Water	0.185	13.900	13.715	01/26/2018	10:03
3	CAL3	10	1	Water	0.318	29.700	29.382	01/26/2018	10:10
4	CAL4	15	1	Water	0.178	46.400	46.222	01/26/2018	10:17
5	CAL5	20	1	Water	0.178	59.700	59.522	01/26/2018	10:25
6	CAL6	25	1	Water	0.180	78.200	78.020	01/26/2018	10:32
7	CAL7	30	1	Water	0.182	92.700	92.518	01/26/2018	10:38
8	CAL8	35	1	Water	0.196	110.900	110.704	01/26/2018	10:45
9	CAL9	40	1	Water	0.200	127.000	126.800	01/26/2018	10:51



Analytical Summary Report

Analysis Method: 9038
Parameter: Sulfate
Run Number: LB92903

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	Initial Weight	Final Vol	True Value	DF	MATRIX	Initial Reading	Final Reading	Difference	Result (mg/L)	AnalDate	Anal Time
1	ICV			20	1	Water	0.189	59.600	59.411	19.219	01/26/2018	10:56
2	ICB				1	Water	0.199	0.289	0.090	0.593	01/26/2018	11:02
3	CCV1			20	1	Water	0.190	60.200	60.010	19.407	01/26/2018	12:09
4	CCB1				1	Water	0.186	0.280	0.094	0.595	01/26/2018	12:15
5	LB92903BL	50	50		1	Water	0.186	0.307	0.121	0.603	01/26/2018	12:27
6	LB92903BS	50	50	20	1	Water	0.193	58.900	58.707	18.998	01/26/2018	12:35
7	J1161-07	50	50		1	Water	0.209	1.860	1.651	1.083	01/26/2018	12:45
8	J1161-08	50	50		1	Water	0.212	25.400	25.188	8.474	01/26/2018	12:53
9	J1161-09	50	50		1	Water	0.175	2.330	2.155	1.242	01/26/2018	13:01
10	CCV2			20	1	Water	0.192	60.100	59.908	19.375	01/26/2018	13:08
11	CCB2				1	Water	0.182	0.286	0.104	0.598	01/26/2018	13:15