


ISM02.4

LB 92910
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45856



A= 0.0000e+000

B= 2.1835e-004

C= -2.0329e-002

Rho= 0.9997665

Accept=Accepted

slope
4.1835e-004

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.00	0.000	-0.012	-0.012	39	0.000	39					-
0.2	0.200	0.207	0.007	1039	0.0 %	1039					3
2.5	2.500	2.589	0.089	11951	0.0 %	11951					4
5.0	5.000	4.846	-0.154	22285	0.0 %	22285					-3
7.5	7.500	7.578	0.078	34800	0.0 %	34800					1
10.0	10.000	9.992	-0.008	45856	0.0 %	45856					0

NB
1926118

LB92910
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	39	-		0 ISM02.4	PPB	1/26/2018 17:50	Std
	0.2 S0.2	1039	-		0.2 ISM02.4	PPB	1/26/2018 17:52	Std
	2.5 S2.5	11951	-		2.5 ISM02.4	PPB	1/26/2018 17:54	Std
	5 S5	22285	-		5 ISM02.4	PPB	1/26/2018 17:57	Std
	7.5 S7.5	34800	-		7.5 ISM02.4	PPB	1/26/2018 17:59	Std
	10 S10	45856	-		10 ISM02.4	PPB	1/26/2018 18:04	Std
ICV	ICV	17924	3.8934 -		ISM02.4	PPB	1/26/2018 18:07	SMPL
ICB	ICB	-106	-0.0435 -		ISM02.4	PPB	1/26/2018 18:09	SMPL
CCV089	CCV089	22210	4.8292 -		ISM02.4	PPB	1/26/2018 18:11	SMPL
CCB089	CCB089	-75	-0.0367 -		ISM02.4	PPB	1/26/2018 18:13	SMPL
PB106036BL	PBS004	163	0.0153 -		ISM02.4	PPB	1/26/2018 18:18	SMPL
J1280-01	MF6B52	461	0.0803 -		ISM02.4	PPB	1/26/2018 18:21	SMPL
J1280-02	MF6B53	191	0.0214 -		ISM02.4	PPB	1/26/2018 18:23	SMPL
J1280-03	MF6B54	306	0.0465 -		ISM02.4	PPB	1/26/2018 18:25	SMPL
J1280-04	MF6B55	300	0.0452 -		ISM02.4	PPB	1/26/2018 18:27	SMPL
J1280-05	MF6B56	411	0.0694 -		ISM02.4	PPB	1/26/2018 18:30	SMPL
J1280-06	MF6B56D	417	0.0707 -		ISM02.4	PPB	1/26/2018 18:32	SMPL
J1280-07	MF6B56S	12048	2.6103 -		ISM02.4	PPB	1/26/2018 18:34	SMPL
J1280-08	MF6B37	191	0.0214 -		ISM02.4	PPB	1/26/2018 18:36	SMPL
J1280-09	MF6B38	366	0.0596 -		ISM02.4	PPB	1/26/2018 18:39	SMPL
J1280-10	MF6B39	348	0.0557 -		ISM02.4	PPB	1/26/2018 18:41	SMPL
J1280-11	MF6B40	568	0.1037 -		ISM02.4	PPB	1/26/2018 18:43	SMPL
J1280-12	MF6B41	767	0.1471 -		ISM02.4	PPB	1/26/2018 18:46	SMPL
J1280-13	MF6B42	535	0.0965 -		ISM02.4	PPB	1/26/2018 18:48	SMPL
J1280-14	MF6B43	409	0.069 -		ISM02.4	PPB	1/26/2018 18:50	SMPL
J1280-15	MF6B44	272	0.0391 -		ISM02.4	PPB	1/26/2018 18:52	SMPL
J1280-16	MF6B45	6433	1.3843 -		ISM02.4	PPB	1/26/2018 18:55	SMPL
J1280-17	MF6B46	495	0.0878 -		ISM02.4	PPB	1/26/2018 18:57	SMPL
PB106037BL	PBS004	112	0.0041 -		ISM02.4	PPB	1/26/2018 18:59	SMPL
J1293-01	MF6B94	1572	0.3229 -		ISM02.4	PPB	1/26/2018 19:01	SMPL
CCV090	CCV090	22244	4.8366 -		ISM02.4	PPB	1/26/2018 19:04	SMPL
CCB090	CCB090	-112	-0.0448 -		ISM02.4	PPB	1/26/2018 19:06	SMPL
J1293-02	MF6B95	605	0.1118 -		ISM02.4	PPB	1/26/2018 19:08	SMPL
J1293-03	MF6B96	631	0.1174 -		ISM02.4	PPB	1/26/2018 19:11	SMPL
J1293-04	MF6B96D	608	0.1124 -		ISM02.4	PPB	1/26/2018 19:13	SMPL
J1293-05	MF6B96S	12802	2.775 -		ISM02.4	PPB	1/26/2018 19:15	SMPL
J1293-06	MF6B97	223	0.0284 -		ISM02.4	PPB	1/26/2018 19:17	SMPL
J1293-07	MF6B98	458	0.0797 -		ISM02.4	PPB	1/26/2018 19:20	SMPL
J1293-08	MF6B99	414	0.0701 -		ISM02.4	PPB	1/26/2018 19:22	SMPL
J1293-09	MF6C00	256	0.0356 -		ISM02.4	PPB	1/26/2018 19:24	SMPL
J1293-10	MF6C01	715	0.1358 -		ISM02.4	PPB	1/26/2018 19:27	SMPL
J1293-11	MF6C02	370	0.0605 -		ISM02.4	PPB	1/26/2018 19:29	SMPL
J1293-12	MF6C03	471	0.0825 -		ISM02.4	PPB	1/26/2018 19:31	SMPL
J1293-13	MF6C04	472	0.0827 -		ISM02.4	PPB	1/26/2018 19:33	SMPL
J1293-14	MF6C05	401	0.0672 -		ISM02.4	PPB	1/26/2018 19:36	SMPL
J1293-15	MF6C06	385	0.0637 -		ISM02.4	PPB	1/26/2018 19:38	SMPL

LB92910
INSTRUMENT ID : CV1

J1293-16	MF6C07	425	0.0725 -	ISM02.4	PPB	1/26/2018 19:40	SMPL
J1293-17	MF6C08	331	0.0519 -	ISM02.4	PPB	1/26/2018 19:42	SMPL
J1293-18	MF6C09	269	0.0384 -	ISM02.4	PPB	1/26/2018 19:45	SMPL
J1293-19	MF6C10	286	0.0421 -	ISM02.4	PPB	1/26/2018 19:47	SMPL
J1293-20	MF6C11	930	0.1827 -	ISM02.4	PPB	1/26/2018 19:49	SMPL
J1293-21	MF6C12	794	0.153 -	ISM02.4	PPB	1/26/2018 19:52	SMPL
CCV091	CCV091	21947	4.7718 -	ISM02.4	PPB	1/26/2018 19:54	SMPL
CCB091	CCB091	-93	-0.0406 -	ISM02.4	PPB	1/26/2018 19:56	SMPL
J1293-22	MF6C13	452	0.0784 -	ISM02.4	PPB	1/26/2018 19:58	SMPL
PB106038BL	PBS004	140	0.0102 -	ISM02.4	PPB	1/26/2018 20:01	SMPL
J1294-01	MF6C14	473	0.083 -	ISM02.4	PPB	1/26/2018 20:03	SMPL
J1294-02	MF6C15	372	0.0609 -	ISM02.4	PPB	1/26/2018 20:05	SMPL
J1294-03	MF6C16	254	0.0351 -	ISM02.4	PPB	1/26/2018 20:08	SMPL
J1294-04	MF6C16D	521	0.0934 -	ISM02.4	PPB	1/26/2018 20:10	SMPL
J1294-05	MF6C16S	12328	2.6715 -	ISM02.4	PPB	1/26/2018 20:12	SMPL
J1294-06	MF6C17	519	0.093 -	ISM02.4	PPB	1/26/2018 20:14	SMPL
J1294-07	MF6C18	733	0.1397 -	ISM02.4	PPB	1/26/2018 20:17	SMPL
J1294-08	MF6C19	907	0.1777 -	ISM02.4	PPB	1/26/2018 20:19	SMPL
J1294-09	MF6C20	839	0.1629 -	ISM02.4	PPB	1/26/2018 20:21	SMPL
J1294-10	MF6C21	502	0.0893 -	ISM02.4	PPB	1/26/2018 20:24	SMPL
J1294-11	MF6C22	638	0.119 -	ISM02.4	PPB	1/26/2018 20:26	SMPL
J1294-12	MF6C23	765	0.1467 -	ISM02.4	PPB	1/26/2018 20:28	SMPL
J1294-13	MF6C24	382	0.0631 -	ISM02.4	PPB	1/26/2018 20:30	SMPL
J1294-14	MF6C25	396	0.0661 -	ISM02.4	PPB	1/26/2018 20:33	SMPL
J1294-15	MF6C26	446	0.0771 -	ISM02.4	PPB	1/26/2018 20:35	SMPL
J1294-16	MF6C27	979	0.1934 -	ISM02.4	PPB	1/26/2018 20:37	SMPL
J1294-17	MF6C28	542	0.098 -	ISM02.4	PPB	1/26/2018 20:39	SMPL
J1294-18	MF6C29	503	0.0895 -	ISM02.4	PPB	1/26/2018 20:42	SMPL
CCV092	CCV092	22152	4.8165 -	ISM02.4	PPB	1/26/2018 20:44	SMPL
CCB092	CCB092	-96	-0.0413 -	ISM02.4	PPB	1/26/2018 20:46	SMPL
J1294-19	MF6C30	545	0.0987 -	ISM02.4	PPB	1/26/2018 20:49	SMPL
J1294-20	MF6C31	698	0.1321 -	ISM02.4	PPB	1/26/2018 20:51	SMPL
J1294-21	MF6C32	554	0.1006 -	ISM02.4	PPB	1/26/2018 20:53	SMPL
J1294-22	MF6C33	470	0.0823 -	ISM02.4	PPB	1/26/2018 20:55	SMPL
PB106039BL	PBS004	30	-0.0138 -	ISM02.4	PPB	1/26/2018 20:58	SMPL
J1313-01	MF6C37	424	0.0723 -	ISM02.4	PPB	1/26/2018 21:00	SMPL
J1313-02	MF6C38	281	0.041 -	ISM02.4	PPB	1/26/2018 21:02	SMPL
J1313-03	MF6C39	963	0.1899 -	ISM02.4	PPB	1/26/2018 21:05	SMPL
J1313-04	MF6C40	132	0.0085 -	ISM02.4	PPB	1/26/2018 21:07	SMPL
J1313-05	MF6C41	111	0.0039 -	ISM02.4	PPB	1/26/2018 21:09	SMPL
J1313-06	MF6C42	112	0.0041 -	ISM02.4	PPB	1/26/2018 21:11	SMPL
J1313-07	MF6C43	870	0.1696 -	ISM02.4	PPB	1/26/2018 21:14	SMPL
J1313-08	MF6C44	1427	0.2913 -	ISM02.4	PPB	1/26/2018 21:16	SMPL
J1313-09	MF6C45	760	0.1456 -	ISM02.4	PPB	1/26/2018 21:18	SMPL
J1313-10	MF6C46	99	0.0013 -	ISM02.4	PPB	1/26/2018 21:21	SMPL
J1313-11	MF6C47	336	0.053 -	ISM02.4	PPB	1/26/2018 21:23	SMPL
J1313-12	MF6C48	166	0.0159 -	ISM02.4	PPB	1/26/2018 21:25	SMPL

LB92910

INSTRUMENT ID : CV1

J1313-13	MF6C49	427	0.0729 -	ISM02.4	PPB	1/26/2018 21:27 SMPL
J1313-14	MF6C50	615	0.114 -	ISM02.4	PPB	1/26/2018 21:30 SMPL
J1313-15	MF6C51	208	0.0251 -	ISM02.4	PPB	1/26/2018 21:32 SMPL
CCV093	CCV093	21959	4.7744 -	ISM02.4	PPB	1/26/2018 21:34 SMPL
CCB093	CCB093	-26	-0.026 -	ISM02.4	PPB	1/26/2018 21:37 SMPL
J1313-16	MF6C52	228	0.0295 -	ISM02.4	PPB	1/26/2018 21:39 SMPL
J1313-17	MF6C53	157	0.014 -	ISM02.4	PPB	1/26/2018 21:41 SMPL
J1313-18	MF6C54	437	0.0751 -	ISM02.4	PPB	1/26/2018 21:43 SMPL
J1313-19	MF6C55	180	0.019 -	ISM02.4	PPB	1/26/2018 21:46 SMPL
J1313-20	MF6C56	366	0.0596 -	ISM02.4	PPB	1/26/2018 21:48 SMPL
J1313-21	MF6C56D	416	0.0705 -	ISM02.4	PPB	1/26/2018 21:50 SMPL
J1313-22	MF6C56S	12266	2.6579 -	ISM02.4	PPB	1/26/2018 21:53 SMPL
CCV094	CCV094	21838	4.748 -	ISM02.4	PPB	1/26/2018 21:57 SMPL
CCB094	CCB094	-24	-0.0256 -	ISM02.4	PPB	1/26/2018 21:59 SMPL