



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92923

Temp.: 22.1 °C

BalanceID: WC SC-4

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 2.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	01/26/2018	17:00
2	CAL2	1	Water			7.01	01/26/2018	17:01
3	CAL3	1	Water			10.03	01/26/2018	17:02
4	ICV	1	Water			7.00	01/26/2018	17:03
5	CCV1	1	Water			2.01	01/26/2018	17:05
6	J1022-11	1	Solid	20.02	20	9.13	01/26/2018	17:06
7	J1022-11DUP	1	Solid	20.03	20	9.14	01/25/2018	17:08
8	J1338-02	1	Solid	20.04	20	5.65	01/26/2018	17:10
9	J1340-01	1	Solid	20.02	20	6.77	01/26/2018	17:11
10	J1340-02	1	Solid	20.03	20	7.98	01/26/2018	17:13
11	J1341-01	1	Solid	20.02	20	8.13	01/25/2018	17:15
12	J1341-02	1	Solid	20.01	20	8.62	01/26/2018	17:18
13	J1341-03	1	Solid	20.03	20	7.89	01/26/2018	17:20
14	J1345-01	1	Solid	20.02	20	7.81	01/26/2018	17:22
15	J1345-02	1	Solid	20.03	20	7.45	01/26/2018	17:25
16	CCV2	1	Water			12.02	01/26/2018	17:26
17	J1345-03	1	Solid	20.01	20	7.88	01/26/2018	17:29
18	J1345-04	1	Solid	20.04	20	8.03	01/26/2018	17:30
19	J1346-04	1	Solid	20.05	20	9.75	01/26/2018	17:31
20	CCV3	1	Water			2.01	01/26/2018	17:33

LB 92923

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I1346

WorkList ID : 107982

Date : 1/26/2018 4:23:04 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/26/2018	Solid	J1022-11	Corrosivity	Cool 4 deg C	PSEG05		SU-03-012618	01/26/2018	9045C
01/26/2018	Solid	J1338-02	Corrosivity	Cool 4 deg C	PSEG02	M52	OILY-STONES	01/26/2018	9045C
01/25/2018	Solid	J1340-01	Corrosivity	Cool 4 deg C	CLEA11	N23	CHASSIS-WASH	01/25/2018	9045C
01/25/2018	Solid	J1340-02	Corrosivity	Cool 4 deg C	CLEA11	N23	BUS-WASH	01/25/2018	9045C
01/24/2018	Solid	J1341-01	Corrosivity	Cool 4 deg C	CLEA11		CHASSIS-WASH	01/24/2018	9045C
01/24/2018	Solid	J1341-02	Corrosivity	Cool 4 deg C	CLEA11		BUS-WASH	01/24/2018	9045C
01/24/2018	Solid	J1341-03	Corrosivity	Cool 4 deg C	CLEA11		RECLAIMER-SLUDGE-PIT	01/24/2018	9045C
01/24/2018	Solid	J1345-01	Corrosivity	Cool 4 deg C	CLEA11	N23	BUS-WASH	01/24/2018	9045C
01/24/2018	Solid	J1345-02	Corrosivity	Cool 4 deg C	CLEA11	N23	RECLAIMER-TANK	01/24/2018	9045C
01/24/2018	Solid	J1345-03	Corrosivity	Cool 4 deg C	CLEA11	N23	ZIPPER-DRAIN	01/24/2018	9045C
01/24/2018	Solid	J1345-04	Corrosivity	Cool 4 deg C	CLEA11	N23	FLOOR-DRAIN	01/24/2018	9045C
01/26/2018	Solid	J1346-04	Corrosivity	Cool 4 deg C	PSEG01	M52	FW-1-2	01/26/2018	9045C

Date/Time

Received by:

Relinquished by:

Date/Time

Received by:

Relinquished by: