


245.1

LB92941
INSTRUMENT ID: CYL

Linear

 μ Abs.:

36907



A= 0.0000e+000

B= 2.7241e-004

C= 1.1628e-002

Rho= 0.9998792

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	-19	0.000	-19				
0.05	0.050	0.056	0.006	164	0.0 %	164				
0.2	0.200	0.222	0.022	774	0.0 %	774				
2.5	2.500	2.456	-0.044	8975	0.0 %	8975				
5.0	5.000	5.058	0.058	18525	0.0 %	18525				
7.5	7.500	7.385	-0.115	27068	0.0 %	27068				
10.0	10.000	10.065	0.065	36907	0.0 %	36907				

LB92941 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 27 Jan 2018 16:18:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-19	-	0.0000	245.1	PPB	27 Jan 2018 16:20:37	S	Std
0.05 - 1	50.05	164	-	0.0500	245.1	PPB	27 Jan 2018 16:22:45	S	Std
0.2 - 1	50.2	774	-	0.2000	245.1	PPB	27 Jan 2018 16:24:53	S	Std
2.5 - 1	50.5	8975	-	2.5000	245.1	PPB	27 Jan 2018 16:27:01	S	Std
5.0 - 1	55	18525	-	5.0000	245.1	PPB	27 Jan 2018 16:29:09	S	Std
7.5 - 1	57.5	27068	-	7.5000	245.1	PPB	27 Jan 2018 16:34:38	S	Std
10.0 - 1	510	36907	-	10.0000	245.1	PPB	27 Jan 2018 16:39:12	S	Std
ICV69 - 1	ICV69	14040	3.8362	-245.1		PPB	27 Jan 2018 16:41:51	U	SMPL
ICB69 - 1	ICB69	-123	-0.0219	-245.1		PPB	27 Jan 2018 16:43:57	U	SMPL
CCV10 - 1	CCV10	18505	5.0525	-245.1		PPB	27 Jan 2018 16:46:03	U	SMPL
CCB10 - 1	CCB10	-154	-0.0303	-245.1		PPB	27 Jan 2018 16:48:10	U	SMPL
CRA - 1	CRA	762	0.2192	-245.1		PPB	27 Jan 2018 16:50:17	U	SMPL
HighStd - 1	HighStd	36863	10.0533	-245.1		PPB	27 Jan 2018 16:52:23	U	SMPL
ChkStd - 1	ChkStd	24532	6.6943	-245.1		PPB	27 Jan 2018 16:54:30	U	SMPL
PB106112BL - 1	PBW	-112	-0.0189	-245.1		PPB	27 Jan 2018 16:59:16	U	SMPL
PB106112BS - 1	LCSW	13203	3.6082	-245.1		PPB	27 Jan 2018 17:01:23	U	SMPL
J1190-05 - 1	PT-HG-WS	19531	5.3320	-245.1		PPB	27 Jan 2018 17:03:31	U	SMPL
CCV11 - 1	CCV11	18513	5.0547	-245.1		PPB	27 Jan 2018 17:14:09	U	SMPL
CCB11 - 1	CCB11	-131	-0.0241	-245.1		PPB	27 Jan 2018 17:16:15	U	SMPL