

LB92942


245.1

INSTRUMENT ID: CV4

Linear

 μ Abs.:

36112



A= 0.0000e+000

B= 2.7557e-004

C= -6.6764e-003

Rho= 0.9999562

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	-13	0.000	-13				
0.05	0.050	0.042	-0.008	176	0.0 %	176				
0.2	0.200	0.201	0.001	752	0.0 %	752				
2.5	2.500	2.477	-0.023	9012	0.0 %	9012				
5.0	5.000	5.048	0.048	18343	0.0 %	18343				
7.5	7.500	7.549	0.049	27417	0.0 %	27417				
10.0	10.000	9.945	-0.055	36112	0.0 %	36112				

LB92942 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 27 Jan 2018 17:42:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-13	-	0.0000	245.1	PPB	27 Jan 2018 17:43:43	S	Std
0.05 - 1	50.05	176	-	0.0500	245.1	PPB	27 Jan 2018 17:45:52	S	Std
0.2 - 1	50.2	752	-	0.2000	245.1	PPB	27 Jan 2018 17:48:00	S	Std
2.5 - 1	52.5	9012	-	2.5000	245.1	PPB	27 Jan 2018 17:50:08	S	Std
5.0 - 1	55	18343	-	5.0000	245.1	PPB	27 Jan 2018 17:52:17	S	Std
7.5 - 1	57.5	27417	-	7.5000	245.1	PPB	27 Jan 2018 17:56:48	S	Std
10.0 - 1	510	36112	-	10.0000	245.1	PPB	27 Jan 2018 17:58:55	S	Std
ICV70 - 1	ICV70	13965	3.8416	-245.1		PPB	27 Jan 2018 18:01:36	U	SMPL
ICB70 - 1	ICB70	-127	-0.0417	-245.1		PPB	27 Jan 2018 18:03:42	U	SMPL
CCV13 - 1	CCV13	18371	5.0558	-245.1		PPB	27 Jan 2018 18:05:49	U	SMPL
CCB13 - 1	CCB13	-136	-0.0442	-245.1		PPB	27 Jan 2018 18:07:56	U	SMPL
CRA - 1	CRA	739	0.1970	-245.1		PPB	27 Jan 2018 18:10:03	U	SMPL
HighStd - 1	HighStd	36589	10.0760	-245.1		PPB	27 Jan 2018 18:12:09	U	SMPL
ChkStd - 1	ChkStd	24481	6.7395	-245.1		PPB	27 Jan 2018 18:14:16	U	SMPL
PB106116BL - 1	PBW	-56	-0.0221	-245.1		PPB	27 Jan 2018 18:19:54	U	SMPL
PB106116BS - 1	LCSW	13329	3.6664	-245.1		PPB	27 Jan 2018 18:22:01	U	SMPL
J1193-04 - 1	1011B	94182	25.9467	-245.1		PPB	27 Jan 2018 18:24:09	U	SMPL
CCV14 - 1	CCV14	18234	5.0180	-245.1		PPB	27 Jan 2018 18:31:47	U	SMPL
CCB14 - 1	CCB14	-212	-0.0651	-245.1		PPB	27 Jan 2018 18:33:57	U	SMPL
J1193-04DLX5 - 1	1011B	18918	5.2065	-245.1		PPB	27 Jan 2018 18:38:11	U	SMPL
CCV15 - 1	CCV15	18409	5.0662	-245.1		PPB	27 Jan 2018 18:42:26	U	SMPL
CCB15 - 1	CCB15	-154	-0.0491	-245.1		PPB	27 Jan 2018 18:44:33	U	SMPL