


7470A

LB 92943
INSTRUMENT ID: CV2

Linear

 μ Abs.:

36526



A= 0.0000e+000

B= 2.7480e-004

C= 8.5340e-003

Rho= 0.9997846

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	-45	0.000	-45				
0.2	0.200	0.213	0.013	744	0.0 %	744				
2.5	2.500	2.458	-0.042	8912	0.0 %	8912				
5.0	5.000	5.118	0.118	18595	0.0 %	18595				
7.5	7.500	7.369	-0.131	26786	0.0 %	26786				
10.0	10.000	10.046	0.046	36526	0.0 %	36526				

LB92943 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Jan 2018 18:52:16

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	-45	-	0.0000	7470A	PPB	27 Jan 2018 18:53:09	S	Std
0.2 - 1	SO.2	744	-	0.2000	7470A	PPB	27 Jan 2018 18:55:17	S	Std
2.5 - 1	SO.5	8912	-	2.5000	7470A	PPB	27 Jan 2018 18:57:26	S	Std
5.0 - 1	SO	18595	-	5.0000	7470A	PPB	27 Jan 2018 18:59:34	S	Std
7.5 - 1	SO.5	26786	-	7.5000	7470A	PPB	27 Jan 2018 19:01:42	S	Std
10.0 - 1	SO	36526	-	10.0000	7470A	PPB	27 Jan 2018 19:03:49	S	Std
ICV71 - 1	ICV71	13528	3.7260	-	7470A	PPB	27 Jan 2018 19:06:23	U	SMPL
ICB71 - 1	ICB71	-163	-0.0363	-	7470A	PPB	27 Jan 2018 19:08:29	U	SMPL
CCV16 - 1	CCV16	18895	5.2008	-	7470A	PPB	27 Jan 2018 19:10:36	U	SMPL
CCB16 - 1	CCB16	-138	-0.0294	-	7470A	PPB	27 Jan 2018 19:12:42	U	SMPL
CRA - 1	CRA	728	0.2086	-	7470A	PPB	27 Jan 2018 19:14:49	U	SMPL
HighStd - 1	HighStd	36496	10.0375	-	7470A	PPB	27 Jan 2018 19:16:55	U	SMPL
ChkStd - 1	ChkStd	25351	6.9749	-	7470A	PPB	27 Jan 2018 19:19:02	U	SMPL
PB106117BL - 1	PBW	-25	0.0017	-	7470A	PPB	27 Jan 2018 19:23:34	U	SMPL
PB106117BS - 1	LCSW	13708	3.7754	-	7470A	PPB	27 Jan 2018 19:25:41	U	SMPL
J1195-05 - 1	1011B	92075	25.3104	-	7470A	PPB	27 Jan 2018 19:27:48	U	SMPL
J1195-05DLX5 - 1	1011B	17928	4.9351	-	7470A	PPB	27 Jan 2018 19:38:36	U	SMPL
CCV17 - 1	CCV17	18181	5.0046	-	7470A	PPB	27 Jan 2018 19:40:43	U	SMPL
CCB17 - 1	CCB17	-142	-0.0305	-	7470A	PPB	27 Jan 2018 19:42:51	U	SMPL