


7471B

LB92944
INSTRUMENT ID: CV2

Linear

 μ Abs.:

43213



A= 0.0000e+000

B= 2.3064e-004

C= -1.1394e-002

Rho= 0.9997841

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.013	-0.013	-5	0.000	-5				
0.2	0.200	0.206	0.006	943	0.0 %	943	0			
2.5	2.500	2.557	0.057	11136	0.0 %	11136				
5.0	5.000	4.875	-0.125	21187	0.0 %	21187				
7.5	7.500	7.619	0.119	33085	0.0 %	33085				
10.0	10.000	9.955	-0.045	43213	0.0 %	43213				

LB92944 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Jan 2018 19:50:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-5	-	0.00007471B	PPB		27 Jan 2018 19:51:36	S	Std
0.2 - 1	50.2	943	-	0.20007471B	PPB		27 Jan 2018 19:53:53	S	Std
2.5 - 1	52.5	11136	-	2.50007471B	PPB		27 Jan 2018 19:56:11	S	Std
5.0 - 1	55	21187	-	5.00007471B	PPB		27 Jan 2018 19:58:28	S	Std
7.5 - 1	57.5	33085	-	7.50007471B	PPB		27 Jan 2018 20:00:45	S	Std
10.0 - 1	51.0	43213	-	10.00007471B	PPB		27 Jan 2018 20:03:02	S	Std
ICV72 - 1	ICV72	17208	3.9574	-7471B	PPB		27 Jan 2018 20:05:46	U	SMPL
ICB72 - 1	ICB72	-75	-0.0287	-7471B	PPB		27 Jan 2018 20:08:02	U	SMPL
CCV18 - 1	CCV18	20979	4.8271	-7471B	PPB		27 Jan 2018 20:10:18	U	SMPL
CCB18 - 1	CCB18	-89	-0.0319	-7471B	PPB		27 Jan 2018 20:12:34	U	SMPL
CRA - 1	CRA	963	0.2107	-7471B	PPB		27 Jan 2018 20:14:50	U	SMPL
HighStd - 1	HighStd	43781	10.0861	-7471B	PPB		27 Jan 2018 20:17:06	U	SMPL
ChkStd - 1	ChkStd	30225	6.9596	-7471B	PPB		27 Jan 2018 20:19:22	U	SMPL
PB106114BL - 1	PBS	-177	-0.0522	-7471B	PPB		27 Jan 2018 20:21:40	U	SMPL
PB106114BS - 1	LCSS	14577	3.3506	-7471B	PPB		27 Jan 2018 20:23:58	U	SMPL
J1022-10 - 1	SU-03-012618	2529	0.5719	-7471B	PPB		27 Jan 2018 20:26:15	U	SMPL
J1195-01 - 1	1022	640102	HIGH	-7471B	PPB		27 Jan 2018 20:28:33	X	
J1338-01 - 1	OILY-STONES	954	0.2086	-7471B	PPB		27 Jan 2018 20:33:26	U	SMPL
J1344-01 - 1	8-SAN-2-EW(5-6)	21391	4.9222	-7471B	PPB		27 Jan 2018 20:35:44	U	SMPL
J1344-02 - 1	8-SAN-2-WW(5-6)	3515	0.7993	-7471B	PPB		27 Jan 2018 20:37:59	U	SMPL
CCV19 - 1	CCV19	21055	4.8447	-7471B	PPB		27 Jan 2018 20:40:16	U	SMPL
CCB19 - 1	CCB19	-74	-0.0285	-7471B	PPB		27 Jan 2018 20:42:31	U	SMPL
J1344-03 - 1	8-SAN-3-EW(5-6)	1529	0.3412	-7471B	PPB		27 Jan 2018 20:44:48	U	SMPL
J1344-04 - 1	8-SAN-3-WW(5-6)	1485	0.3311	-7471B	PPB		27 Jan 2018 20:47:04	U	SMPL
J1344-05 - 1	8-SAN-4-WW(4-5)	958	0.2096	-7471B	PPB		27 Jan 2018 20:49:20	U	SMPL
J1344-06 - 1	8-SAN-4-EW(4-5)	877	0.1909	-7471B	PPB		27 Jan 2018 20:51:37	U	SMPL
J1344-06DUP - 1	8-SAN-4-EW(4-5)DUP	1024	0.2248	-7471B	PPB		27 Jan 2018 20:53:53	U	SMPL
J1344-06MS - 1	8-SAN-4-EW(4-5)MS	15754	3.6221	-7471B	PPB		27 Jan 2018 20:56:10	U	SMPL
J1344-06MSD - 1	8-SAN-4-EW(4-5)MSD	16226	3.7309	-7471B	PPB		27 Jan 2018 20:58:27	U	SMPL
J1344-07 - 1	8-SAN-4-EW(7-8)	1088	0.2395	-7471B	PPB		27 Jan 2018 21:00:45	U	SMPL
J1344-06LX5 - 1		176	0.0292	-7471B	PPB		27 Jan 2018 21:03:02	U	SMPL
J1344-06A - 1		17251	3.9673	-7471B	PPB		27 Jan 2018 21:05:18	U	SMPL
CCV20 - 1	CCV20	21097	4.8544	-7471B	PPB		27 Jan 2018 21:07:34	U	SMPL
CCB20 - 1	CCB20	-69	-0.0273	-7471B	PPB		27 Jan 2018 21:09:50	U	SMPL
J1195-01DLX10 - 1	1022	103968	23.9675	-7471B	PPB		27 Jan 2018 21:14:23	U	SMPL
J1195-01DL2X50 - 1	1022	19502	4.4865	-7471B	PPB		27 Jan 2018 21:30:52	U	SMPL
CCV21 - 1	CCV21	20985	4.8285	-7471B	PPB		27 Jan 2018 21:33:09	U	SMPL
CCB21 - 1	CCB21	-72	-0.0280	-7471B	PPB		27 Jan 2018 21:35:26	U	SMPL