



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB92956
Analysis Date: 01/29/2018
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: Sweta
REVIEWED BY: apatel
Extraction Date: 01/29/2018
Extraction IN Time: 09:30
Extraction OUT Time: 10:30

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB92956BL	LB92956BL	WATER	2	1000	100	3.1095	3.1095	3.02	3.1095	3.1095	0	0
2	LB92956BS	LB92956BS	WATER	2	1000	100	3.1112	3.1112	3.01	3.1293	3.1293	0.0181	18.1
3	J1350-01	STORM-DRAIN-1	WATER	2	1000	100	3.1137	3.1137	3.00	3.1139	3.1139	0.0002	0.2
4	J1350-01DUP	STORM-DRAIN-1DUP	WATER	2	1000	100	3.1154	3.1154	3.02	3.1157	3.1157	0.0003	0.3
5	J1350-02	STORM-DRAIN-2	WATER	2	1000	100	3.0829	3.0829	3.04	3.083	3.083	0.0001	0.1
6	J1350-02MS	STORM-DRAIN-2	WATER	2	1000	100	3.1132	3.1132	3.02	3.1333	3.1333	0.0201	20.1



QC Batch# LB92956 Test: TPH Analysis Date: 01/29/2018

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2303
pH Paper 0-14	W2287
Sodium Sulfate	EP1806
1:1 HCL	WP59249
Silica Gel	W2199
Sand	N/A

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP58006
LCSWD	N/A	N/A
MS/MSD	10.00 ML	WP58005

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0021 In OVEN TEMP : 71 Dessicator Time In : 12:05
1.0000 gram Balance: 1.0005 In Time: 11:35
Bal Check Time: 09:45

After Analysis

0.0020 gram Balance: 0.0018 Out OVEN TEMP: 71 Dessicator Time Out: 13:45
1.0000 gram Balance: 1.0003 Out Time: 13:05
Bal Check Time: 14:05

WORKLIST(Hardcopy Internal Chain)

LB 92 956

Worklist Name : TPH012918

Worklist ID : 108040

Date : 1/29/2018 9:12:40 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/02/2018	Water	J1350-01	TPH	Conc H2SO4 to	KWAR01	N61	STORM-DRAIN-1	01/26/2018	1664A
02/02/2018	Water	J1350-02	TPH	Conc H2SO4 to	KWAR01	N61	STORM-DRAIN-2	01/26/2018	1664A

Date/Time 01/29/18 9:15 AM

Received by: SP

Relinquished by: QP

Date/Time 01/29/18 10:40 AM

Received by: QP

Relinquished by: SP