


ISM02.4

LB 92961
INSTRUMENT ID: CV1

Linear

 μ Abs.:

41998



A= 0.0000e+000

B= 2.4031e-004

C= -6.2694e-002

Rho= 0.9999517

Accept=Accepted

slope
 $y = 1.21e5 x$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.00	0.000	-0.017	-0.017	191	0.000	191					-
0.2	0.200	0.215	0.015	1156	0.0 %	1156					7
2.5	2.500	2.499	-0.001	10659	0.0 %	10659					0
5.0	5.000	5.042	0.042	21241	0.0 %	21241					1
7.5	7.500	7.431	-0.069	31184	0.0 %	31184					1
10.0	10.000	10.030	0.030	41998	0.0 %	41998					0

NB
1/29/18

LB92961
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	191	-		0 ISM02.4	PPB	1/29/2018 16:21	Std
	0.2 S0.2	1156	-		0.2 ISM02.4	PPB	1/29/2018 16:23	Std
	2.5 S2.5	10659	-		2.5 ISM02.4	PPB	1/29/2018 16:25	Std
	5 S5	21241	-		5 ISM02.4	PPB	1/29/2018 16:28	Std
	7.5 S7.5	31184	-		7.5 ISM02.4	PPB	1/29/2018 16:30	Std
	10 S10	41998	-		10 ISM02.4	PPB	1/29/2018 16:32	Std
ICV	ICV	16209	3.8325 -		ISM02.4	PPB	1/29/2018 16:35	SMPL
ICB	ICB	68	-0.0464 -		ISM02.4	PPB	1/29/2018 16:38	SMPL
CCV022	CCV022	21342	5.066 -		ISM02.4	PPB	1/29/2018 16:40	SMPL
CCB022	CCB022	74	-0.0449 -		ISM02.4	PPB	1/29/2018 16:42	SMPL
PB106145BL	PBW003	217	-0.0105 -		ISM02.4	PPB	1/29/2018 16:44	SMPL
J1226-11	MF6D42	-114	-0.0901 -		ISM02.4	PPB	1/29/2018 16:47	SMPL
J1226-12	MF6D43	-111	-0.0894 -		ISM02.4	PPB	1/29/2018 16:49	SMPL
J1226-21	MF6D44	-86	-0.0834 -		ISM02.4	PPB	1/29/2018 16:51	SMPL
PB106146BL	PBW003	176	-0.0204 -		ISM02.4	PPB	1/29/2018 16:56	SMPL
J1323-07	MX2B99	11885	2.7934 -		ISM02.4	PPB	1/29/2018 16:58	SMPL
J1323-08	MX2B99D	11762	2.7639 -		ISM02.4	PPB	1/29/2018 17:00	SMPL
J1323-09	MX2B99S	16162	3.8212 -		ISM02.4	PPB	1/29/2018 17:03	SMPL
CCV023	CCV023	21248	5.0435 -		ISM02.4	PPB	1/29/2018 17:05	SMPL
CCB023	CCB023	56	-0.0492 -		ISM02.4	PPB	1/29/2018 17:07	SMPL