



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB92967

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP61448
COD calibration std. 100 ppm	WP61444
COD calibration std. 10 ppm	WP61443
COD calibration std. 150 ppm	WP61445
COD calibration std. 50 ppm	WP61446
COD calibration std. 0 ppm	WP61442
COD CCV std, 50ppm	WP61449
COD LOD std, 1PPM	WP61451
COD LOQ std, 10.0PPM	WP61450
COD Digestion Vials Low Level 0-150Mg/L	W2291

Temp In (C) : <u>150</u>	Date In: <u>01/30/2018</u>	Time In: <u>09:30</u>
Temp Out (C) : <u>150</u>	Date Out: <u>01/30/2018</u>	Time Out: <u>11:30</u>

Intercept: 0.1033

Slope: 1.0048

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	01/30/2018	12:20
2	CAL2	10	1	Water	10.000	01/30/2018	12:20
3	CAL3	50	1	Water	51.000	01/30/2018	12:21
4	CAL4	100	1	Water	100.000	01/30/2018	12:21
5	CAL5	150	1	Water	151.000	01/30/2018	12:22



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB92967

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.658	01/30/2018	12:22
2	ICB				1	Water	0.000	-0.103	01/30/2018	12:23
3	CCV1	50			1	Water	51.000	50.654	01/30/2018	12:23
4	CCB1				1	Water	0.000	-0.103	01/30/2018	12:24
5	LB92967BL				1	Water	0.000	-0.103	01/30/2018	12:24
6	LB92967BS	50			1	Water	51.000	50.654	01/30/2018	12:25
7	J1161-07					Water	2.000	1.888	01/30/2018	12:25
8	J1161-08				1	Water	8.000	7.859	01/30/2018	12:26
9	J1161-09				1	Water	3.000	2.883	01/30/2018	12:26
10	J1192-18				5	Water	40.000	39.706	01/30/2018	12:27
11	J1350-01				1	Water	10.000	9.849	01/30/2018	12:27
12	J1350-02				1	Water	0.000	-0.103	01/30/2018	12:28
13	J1350-02DUP				1	Water	0.000	-0.103	01/30/2018	12:28
14	J1350-02MS	50			1	Water	50.000	49.658	01/30/2018	12:29
15	CCV2	50			1	Water	51.000	50.654	01/30/2018	12:29
16	CCB2				1	Water	0.000	-0.103	01/30/2018	12:30
17	J1350-02MSD	50			1	Water	50.000	49.658	01/30/2018	12:30
18	CCV3	50			1	Water	51.000	50.654	01/30/2018	12:31
19	CCB3				1	Water	0.000	-0.103	01/30/2018	12:31

WORKLIST(Hardcopy Internal Chain)

LB 92967

WorkList Name : COD012918

WorkList ID : 108042

Date : 1/29/2018 9:59:33 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/09/2018	Water	J1350-01	COD	Conc H2SO4 4c	KWAR01	N61	STORM-DRAIN-1	01/26/2018	SM5220 D
02/09/2018	Water	J1350-02	COD	Conc H2SO4 4c	KWAR01	N61	STORM-DRAIN-2	01/26/2018	SM5220 D

Date/Time 01/30/18 9:00
 Received by: IB
 Relinquished by: CP

Date/Time 01/30/18 9:50
 Received by: CP
 Relinquished by: IB