

ISM02.4


LB92972

INSTRUMENT ID: CV4

Linear

 μ Abs.:

45912



A= 0.0000e+000

B= 2.1930e-004

C= -5.6056e-002

Rho= 0.9999729

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	q/b
0.00	0.000	-0.021	-0.021	162	0.000	162					1
0.2	0.200	0.200	0.000	1169	0.0 %	1169					0
2.5	2.500	2.552	0.052	11894	0.0 %	11894					2
5.0	5.000	4.970	-0.030	22920	0.0 %	22920					1
7.5	7.500	7.485	-0.015	34386	0.0 %	34386					0
10.0	10.000	10.013	0.013	45912	0.0 %	45912					0

NB
1/29/18

LB92972

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0-0	162	-		0 ISM02.4	PPB	1/29/2018 17:37	Std
	0.2 S0.2	1169	-		0.2 ISM02.4	PPB	1/29/2018 17:39	Std
	2.5 S2.5	11894	-		2.5 ISM02.4	PPB	1/29/2018 17:42	Std
	5 S5	22920	-		5 ISM02.4	PPB	1/29/2018 17:44	Std
	7.5 S7.5	34386	-		7.5 ISM02.4	PPB	1/29/2018 17:46	Std
	10 S10	45912	-		10 ISM02.4	PPB	1/29/2018 17:51	Std
ICV	ICV	17771	3.8412 -		ISM02.4	PPB	1/29/2018 17:54	SMPL
ICB	ICB	18	-0.0521 -		ISM02.4	PPB	1/29/2018 17:56	SMPL
CCV024	CCV024	23072	5.0037 -		ISM02.4	PPB	1/29/2018 18:01	SMPL
CCB024	CCB024	10	-0.0539 -		ISM02.4	PPB	1/29/2018 18:03	SMPL
PB106147BL	PBS004	160	-0.021 -		ISM02.4	PPB	1/29/2018 18:06	SMPL
J1323-22	MX2C04	8828	1.88 -		ISM02.4	PPB	1/29/2018 18:08	SMPL
J1323-23	MX2C04D	8154	1.7321 -		ISM02.4	PPB	1/29/2018 18:10	SMPL
J1323-24	MX2C04S	18624	4.0282 -		ISM02.4	PPB	1/29/2018 18:13	SMPL
CCV025	CCV025	22280	4.83 -		ISM02.4	PPB	1/29/2018 18:17	SMPL
CCB025	CCB025	42	-0.0468 -		ISM02.4	PPB	1/29/2018 18:19	SMPL