



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92976

Temp.: 24.6 °C

BalanceID: WC SC-4

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/29/2018	18:00
2	CAL2	1	Water			7.01	01/29/2018	18:01
3	CAL3	1	Water			10.03	01/29/2018	18:02
4	ICV	1	Water			6.99	01/29/2018	18:05
5	CCV1	1	Water			2.01	01/29/2018	18:06
6	J1019-12	1	Solid	20.02	20	7.12	01/29/2018	18:08
7	J1019-12DUP	1	Solid	20.03	20	7.13	01/29/2018	18:10
8	J1019-14	1	Solid	20.03	20	7.57	01/29/2018	18:11
9	J1192-06	1	Solid	20.03	20	10.28	01/29/2018	18:12
10	J1359-01	1	Solid	20.02	20	6.30	01/29/2018	18:14
11	J1359-02	1	Solid	20.01	20	6.91	01/29/2018	18:15
12	J1359-03	1	Solid	20.03	20	7.72	01/29/2018	18:17
13	J1361-02	1	Solid	20.05	20	8.30	01/29/2018	18:19
14	J1363-02	1	Solid	20.02	20	7.74	01/29/2018	18:20
15	J1365-01	1	Solid	20.03	20	8.63	01/29/2018	18:22
16	CCV2	1	Water			12.01	01/29/2018	18:25

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 11365

WorkList ID : 108054

Date : 1/29/2018 5:41:52 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/29/2018	Solid	J1019-12	Corrosivity	Cool 4 deg C	PSEG05		EO-03-012918-A	01/29/2018	9045C
01/29/2018	Solid	J1019-14	Corrosivity	Cool 4 deg C	PSEG05		EO-03-012918-B	01/29/2018	9045C
01/26/2018	Solid	J1359-01	Corrosivity	Cool 4 deg C	CLEA11	M32	BUS-WASH-ZIPPER-DRAINS	01/26/2018	9045C
01/26/2018	Solid	J1359-02	Corrosivity	Cool 4 deg C	CLEA11	M32	RECLAIM-TANK	01/26/2018	9045C
01/26/2018	Solid	J1359-03	Corrosivity	Cool 4 deg C	CLEA11	M32	RECLAIM-SLUDGE-PIT	01/26/2018	9045C
01/29/2018	Solid	J1361-02	Corrosivity	Cool 4 deg C	PSEG02	M52	OILY-STONE	01/29/2018	9045C
01/29/2018	Solid	J1363-02	Corrosivity	Cool 4 deg C	ENTA04	M22	RB-01	01/29/2018	9045C
01/29/2018	Solid	J1365-01	Corrosivity	Cool 4 deg C	SEAC01	M22	SOIL-CUTTINGS	01/29/2018	9045C

Date/Time 01-29-18 5:45 PM
 Received by: JH
 Relinquished by: CR

Date/Time 01-29-18 6:35 PM
 Received by: CR
 Relinquished by: SP