



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: AparnaDate: 01/31/2018DATE & TIME IN: 01/30/2018 16:00OVEN TEMP: 103 °CRun Number: LB93000DATE & TIME OUT: 01/31/2018 10:00OVEN TEMP: 105 °CBalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB93000BL	LB93000BL	76.9626	76.9626	100	76.9626	76.9626	76.9626	0	0
2	J1364-14	MW-03	79.9819	79.9819	100	80.9492	80.9492	80.9492	0.9673	9673
3	J1364-14DUP	MW-03DUP	83.3705	83.3705	100	84.342	84.342	84.342	0.9715	9715
4	J1364-15	MW-04	85.4551	85.4551	100	86.2219	86.2219	86.2219	0.7668	7668
5	J1364-16	MW-05	85.1271	85.1271	100	86.4511	86.4511	86.4511	1.324	13240
6	J1364-17	MW-07	63.8295	63.8295	100	64.2762	64.2762	64.2762	0.4467	4467

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS013018

WorkList ID : 108100

Date : 1/30/2018 3:02:14 PM

LB93000

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/04/2018	Water	J1364-14	TS	Cool 4 deg C	LIRO01	M52	MW-03	01/29/2018	SM2540 B
02/04/2018	Water	J1364-15	TS	Cool 4 deg C	LIRO01	M52	MW-04	01/29/2018	SM2540 B
02/04/2018	Water	J1364-16	TS	Cool 4 deg C	LIRO01	M52	MW-05	01/29/2018	SM2540 B
02/04/2018	Water	J1364-17	TS	Cool 4 deg C	LIRO01	M52	MW-07	01/29/2018	SM2540 B

Date/Time 01/30/18, 15:20

Received by: As

Relinquished by: WJ

Date/Time 01/30/18, 15:50

Received by: WJ

Relinquished by: As