

7471B


LB 93001

INSTRUMENT ID: CVL

Linear

 μ Abs.:

45084



A= 0.0000e+000

B= 2.2210e-004

C= -2.8037e-002

Rho= 0.9999745

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.035	-0.035	-30	0.000	-30				
0.2	0.200	0.195	-0.005	1004	0.0 %	1004	0			
2.5	2.500	2.552	0.052	11617	0.0 %	11617				
5.0	5.000	5.001	0.001	22645	0.0 %	22645				
7.5	7.500	7.501	0.001	33898	0.0 %	33898				
10.0	10.000	9.985	-0.015	45084	0.0 %	45084				

LB93001 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Jan 2018 15:26:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-30	-	0.0000	7471B	PPB	30 Jan 2018 15:35:22	S	Std
0.2 - 1	50.2	1004	-	0.2000	7471B	PPB	30 Jan 2018 15:37:39	S	Std
2.5 - 1	52.5	11617	-	2.5000	7471B	PPB	30 Jan 2018 15:39:57	S	Std
5.0 - 1	55	22645	-	5.0000	7471B	PPB	30 Jan 2018 15:42:14	S	Std
7.5 - 1	57.5	33898	-	7.5000	7471B	PPB	30 Jan 2018 15:44:31	S	Std
10.0 - 1	510	45084	-	10.0000	7471B	PPB	30 Jan 2018 15:46:47	S	Std
ICV74 - 1	ICV74	17512	3.8614	-	7471B	PPB	30 Jan 2018 15:50:05	U	SMPL
ICB74 - 1	ICB74	-96	-0.0494	-	7471B	PPB	30 Jan 2018 15:52:20	U	SMPL
CCV30 - 1	CCV30	22857	5.0486	-	7471B	PPB	30 Jan 2018 15:54:36	U	SMPL
CCB30 - 1	CCB30	-182	-0.0685	-	7471B	PPB	30 Jan 2018 15:56:52	U	SMPL
CRA - 1	CRA	1013	0.1970	-	7471B	PPB	30 Jan 2018 15:59:08	U	SMPL
HighStd - 1	HighStd	45350	10.0444	-	7471B	PPB	30 Jan 2018 16:01:24	U	SMPL
ChkStd - 1	ChkStd	31735	7.0204	-	7471B	PPB	30 Jan 2018 16:03:40	U	SMPL
PB106163BL - 1	PBS	3	-0.0274	-	7471B	PPB	30 Jan 2018 16:08:31	U	SMPL
PB106163BS - 1	LCSS	16377	3.6094	-	7471B	PPB	30 Jan 2018 16:10:47	U	SMPL
J1272-09 - 1	SB9-0.5-1.0	6829	1.4887	-	7471B	PPB	30 Jan 2018 16:13:04	U	SMPL
J1346-01 - 1	FW-1-2	61	-0.0145	-	7471B	PPB	30 Jan 2018 16:15:21	U	SMPL
J1351-01 - 1	DRUM	944	0.1816	-	7471B	PPB	30 Jan 2018 16:17:39	U	SMPL
J1353-01 - 1	EPE-106-5(0-5)	53041	11.7526	-	7471B	PPB	30 Jan 2018 16:19:57	U	SMPL
J1353-02 - 1	EPE-106-5(5-10)	41472	9.1830	-	7471B	PPB	30 Jan 2018 16:22:13	U	SMPL
CCV31 - 1	CCV31	22199	4.9024	-	7471B	PPB	30 Jan 2018 16:24:31	U	SMPL
CCB31 - 1	CCB31	-153	-0.0620	-	7471B	PPB	30 Jan 2018 16:26:48	U	SMPL
J1353-03 - 1	EPE-106-5(15-17)	1914	0.3971	-	7471B	PPB	30 Jan 2018 16:29:05	U	SMPL
J1353-04 - 1	EPE-106-5(20-22)	1156	0.2287	-	7471B	PPB	30 Jan 2018 16:31:20	U	SMPL
J1353-05 - 1	EPE-106-5(50-52)	277	0.0335	-	7471B	PPB	30 Jan 2018 16:33:36	U	SMPL
J1353-06 - 1	EPE-106-5(65-67)	1093	0.2147	-	7471B	PPB	30 Jan 2018 16:35:52	U	SMPL
J1353-07 - 1	EPE-106-5(70-72)	684	0.1239	-	7471B	PPB	30 Jan 2018 16:38:09	U	SMPL
J1353-07DUP - 1	EPE-106-5(70-72)DUP	643	0.1148	-	7471B	PPB	30 Jan 2018 16:40:26	U	SMPL
J1353-07MS - 1	EPE-106-5(70-72)MS	17470	3.8521	-	7471B	PPB	30 Jan 2018 16:42:43	U	SMPL
J1353-07MSD - 1	EPE-106-5(70-72)MSD	17152	3.7815	-	7471B	PPB	30 Jan 2018 16:45:01	U	SMPL
J1353-08 - 1	EPE-106-5(75-77)	605	0.1063	-	7471B	PPB	30 Jan 2018 16:47:18	U	SMPL
J1353-09 - 1	EPE-106-5(80-82)	668	0.1203	-	7471B	PPB	30 Jan 2018 16:49:34	U	SMPL
CCV32 - 1	CCV32	22731	5.0206	-	7471B	PPB	30 Jan 2018 16:51:50	U	SMPL
CCB32 - 1	CCB32	-88	-0.0476	-	7471B	PPB	30 Jan 2018 16:54:06	U	SMPL
J1353-10 - 1	EPE-106-5(85-87)	693	0.1259	-	7471B	PPB	30 Jan 2018 16:56:22	U	SMPL
J1353-11 - 1	EPE-106-5(90-92)	824	0.1550	-	7471B	PPB	30 Jan 2018 16:58:38	U	SMPL
J1360-01 - 1	SPRING-101	1921	0.3986	-	7471B	PPB	30 Jan 2018 17:00:54	U	SMPL
J1361-01 - 1	OILY-STONE	132942	29.4989	-	7471B	PPB	30 Jan 2018 17:03:11	U	SMPL
J1362-01 - 1	WC-1-COMP	6901	1.5047	-	7471B	PPB	30 Jan 2018 17:05:28	U	SMPL
J1362-03 - 1	WC-2-COMP	8180	1.7888	-	7471B	PPB	30 Jan 2018 17:07:49	U	SMPL
J1362-05 - 1	WC-3-COMP	9415	2.0631	-	7471B	PPB	30 Jan 2018 17:10:06	U	SMPL
J1353-07LX5 - 1		49	-0.0172	-	7471B	PPB	30 Jan 2018 17:12:24	U	SMPL
J1353-07A - 1		17688	3.9005	-	7471B	PPB	30 Jan 2018 17:14:41	U	SMPL
J1353-01DLX2 - 1		24767	5.4728	-	7471B	PPB	30 Jan 2018 17:16:57	U	SMPL
CCV33 - 1	CCV33	22530	4.9760	-	7471B	PPB	30 Jan 2018 17:19:13	U	SMPL
CCB33 - 1	CCB33	-181	-0.0682	-	7471B	PPB	30 Jan 2018 17:21:31	U	SMPL
J1361-01DLX5 - 1		24458	5.4042	-	7471B	PPB	30 Jan 2018 17:23:48	U	SMPL
CCV34 - 1	CCV34	22306	4.9262	-	7471B	PPB	30 Jan 2018 17:28:21	U	SMPL
CCV34 - 1	CCV34	-162	-0.0640	-	7471B	PPB	30 Jan 2018 17:30:37	U	SMPL