



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB93006

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	01/30/2018	11:32
2	J1364-14		1	>100.0	>212.0	01/30/2018	11:55
3	J1364-14DUP		1	>100.0	>212.0	01/30/2018	12:20
4	J1364-15		1	>100.0	>212.0	01/30/2018	12:45
5	J1364-16		1	>100.0	>212.0	01/30/2018	13:10
6	J1364-17		1	>100.0	>212.0	01/30/2018	13:35
7	J1365-03		1	>100.0	>212.0	01/30/2018	13:55

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP13018

WorkList ID : 108088

Date : 1/30/2018 10:50:00 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/04/2018	Water	J1364-14 C	Flash Point	Cool 4 deg C	LIRO01	M52	MW-03	01/29/2018	1010A
02/04/2018	Water	J1364-15 C	Flash Point	Cool 4 deg C	LIRO01	M52	MW-04	01/29/2018	1010A
02/04/2018	Water	J1364-16 D	Flash Point	Cool 4 deg C	LIRO01	M52	MW-05	01/29/2018	1010A
02/04/2018	Water	J1364-17 B	Flash Point	Cool 4 deg C	LIRO01	M52	MW-07	01/29/2018	1010A
02/10/2018	Water	J1365-03 ✓	Flash Point	Cool 4 deg C	SEAC01		GROUNDWATER	01/29/2018	1010A

Date/Time 01/30/18 10:55 AM
 Received by: KS
 Relinquished by: CP

Date/Time 01/30/18 14:00
 Received by: CP
 Relinquished by: KS