


7470A

LB93016
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36017



A= 0.0000e+000

B= 2.7802e-004

C= -2.6779e-002

Rho= 0.9999836

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.017	-0.017	34	0.000	34				
0.2	0.200	0.222	0.022	896	0.0 %	896				
2.5	2.500	2.472	-0.028	8989	0.0 %	8989				
5.0	5.000	5.030	0.030	18188	0.0 %	18188				
7.5	7.500	7.506	0.006	27096	0.0 %	27096				
10.0	10.000	9.987	-0.013	36017	0.0 %	36017				

LB93016 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 Jan 2018 10:56:03

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	SO	34	-	0.0000	7470A	PPB	31 Jan 2018 11:02:11	S	Std
0.2 - 1	SO-2	896	-	0.2000	7470A	PPB	31 Jan 2018 11:04:18	S	Std
2.5 - 1	SO-5	8989	-	2.5000	7470A	PPB	31 Jan 2018 11:06:27	S	Std
5.0 - 1	SO	18188	-	5.0000	7470A	PPB	31 Jan 2018 11:08:35	S	Std
7.5 - 1	SO	27096	-	7.5000	7470A	PPB	31 Jan 2018 11:10:42	S	Std
10.0 - 1	SO	36017	-	10.0000	7470A	PPB	31 Jan 2018 11:12:49	S	Std
ICV76 - 1	ICV76	14260	3.9377	-	7470A	PPB	31 Jan 2018 11:15:24	U	SMPL
ICB76 - 1	ICB76	-134	-0.0640	-	7470A	PPB	31 Jan 2018 11:17:30	U	SMPL
CCV35 - 1	CCV35	18139	5.0162	-	7470A	PPB	31 Jan 2018 11:19:37	U	SMPL
CCB35 - 1	CCB35	-116	-0.0590	-	7470A	PPB	31 Jan 2018 11:21:44	U	SMPL
CRA - 1	CRA	873	0.2159	-	7470A	PPB	31 Jan 2018 11:23:51	U	SMPL
HighStd - 1	HighStd	36765	10.1945	-	7470A	PPB	31 Jan 2018 11:25:58	U	SMPL
ChkStd - 1	ChkStd	25736	7.1283	-	7470A	PPB	31 Jan 2018 11:28:04	U	SMPL
PB106194BL - 1	PBW	-192	-0.0802	-	7470A	PPB	31 Jan 2018 11:30:13	U	SMPL
PB106194BS - 1	LCSW	14145	3.9058	-	7470A	PPB	31 Jan 2018 11:32:21	U	SMPL
PB106194TB - 1	PB106166TB	-76	-0.0479	-	7470A	PPB	31 Jan 2018 11:34:29	U	SMPL
J1019-12 - 1	EO-03-012918-A	-1	-0.0271	-	7470A	PPB	31 Jan 2018 11:36:36	U	SMPL
J1019-12DUP - 1	EO-03-012918-ADUP	26	-0.0196	-	7470A	PPB	31 Jan 2018 11:38:45	U	SMPL
J1019-12MS - 1	EO-03-012918-AMS	13557	3.7423	-	7470A	PPB	31 Jan 2018 11:40:54	U	SMPL
J1019-12MSD - 1	EO-03-012918-AMSD	13677	3.7757	-	7470A	PPB	31 Jan 2018 11:43:00	U	SMPL
CCV36 - 1	CCV36	18212	5.0365	-	7470A	PPB	31 Jan 2018 11:45:06	U	SMPL
CCB36 - 1	CCB36	-130	-0.0629	-	7470A	PPB	31 Jan 2018 11:47:13	U	SMPL
J1019-14 - 1	EO-03-012918-B	13	-0.0232	-	7470A	PPB	31 Jan 2018 11:49:21	U	SMPL
J1359-01 - 1	BUS-WASH-ZIPPER-DRAINS	38	-0.0162	-	7470A	PPB	31 Jan 2018 11:51:28	U	SMPL
J1359-02 - 1	RECLAIM-TANK	42	-0.0151	-	7470A	PPB	31 Jan 2018 11:53:34	U	SMPL
J1359-03 - 1	RECLAIM-SLUDGE-PIT	-11	-0.0298	-	7470A	PPB	31 Jan 2018 11:55:41	U	SMPL
J1361-02 - 1	OILY-STONE	23	-0.0204	-	7470A	PPB	31 Jan 2018 11:57:48	U	SMPL
J1363-02 - 1	RB-01	25	-0.0198	-	7470A	PPB	31 Jan 2018 11:59:56	U	SMPL
J1365-01 - 1	SOIL-CUTTINGS	-42	-0.0385	-	7470A	PPB	31 Jan 2018 12:02:04	U	SMPL
PB106197BL - 1	PBW	-121	-0.0604	-	7470A	PPB	31 Jan 2018 12:04:12	U	SMPL
PB106197BS - 1	LCSW	13348	3.6842	-	7470A	PPB	31 Jan 2018 12:06:20	U	SMPL
CCV37 - 1	CCV37	18250	5.0470	-	7470A	PPB	31 Jan 2018 12:10:34	U	SMPL
CCB37 - 1	CCB37	-130	-0.0629	-	7470A	PPB	31 Jan 2018 12:12:41	U	SMPL
J1364-01 - 1	MW-03	3970	1.0769	-	7470A	PPB	31 Jan 2018 12:14:49	U	SMPL
J1364-02 - 1	MW-04	1401	0.3627	-	7470A	PPB	31 Jan 2018 12:16:56	U	SMPL
J1364-03 - 1	MW-05	9896	2.7245	-	7470A	PPB	31 Jan 2018 12:19:02	U	SMPL
J1364-04 - 1	MW-07	677	0.1614	-	7470A	PPB	31 Jan 2018 12:21:09	U	SMPL
J1364-05 - 1	MW-05-DUP	3124	0.8417	-	7470A	PPB	31 Jan 2018 12:23:16	U	SMPL
J1364-06 - 1	EQUIPMENT-BLANK	-234	-0.0918	-	7470A	PPB	31 Jan 2018 12:25:24	U	SMPL
J1364-08 - 1	MW-03	-182	-0.0774	-	7470A	PPB	31 Jan 2018 12:27:32	U	SMPL
J1364-09 - 1	MW-04	-215	-0.0866	-	7470A	PPB	31 Jan 2018 12:29:40	U	SMPL
J1364-10 - 1	MW-05	-224	-0.0891	-	7470A	PPB	31 Jan 2018 12:31:49	U	SMPL
J1364-11 - 1	MW-07	-230	-0.0907	-	7470A	PPB	31 Jan 2018 12:33:55	U	SMPL
CCV38 - 1	CCV38	18160	5.0220	-	7470A	PPB	31 Jan 2018 12:36:02	U	SMPL
CCB38 - 1	CCB38	-94	-0.0529	-	7470A	PPB	31 Jan 2018 12:38:09	U	SMPL
J1364-12 - 1	MW-05-DUP	-193	-0.0804	-	7470A	PPB	31 Jan 2018 12:40:17	U	SMPL
J1364-13 - 1	EQUIPMENT-BLANK	-227	-0.0899	-	7470A	PPB	31 Jan 2018 12:42:24	U	SMPL
J1364-14 - 1	MW-03	3383	0.9138	-	7470A	PPB	31 Jan 2018 12:44:31	U	SMPL
J1364-15 - 1	MW-04	1200	0.3068	-	7470A	PPB	31 Jan 2018 12:46:38	U	SMPL
J1364-16 - 1	MW-05	11110	3.0620	-	7470A	PPB	31 Jan 2018 12:48:45	U	SMPL
J1364-17 - 1	MW-07	387	0.0808	-	7470A	PPB	31 Jan 2018 12:50:53	U	SMPL
J1364-17DUP - 1	MW-07DUP	435	0.0942	-	7470A	PPB	31 Jan 2018 12:53:01	U	SMPL
J1364-17MS - 1	MW-07MS	16920	4.6773	-	7470A	PPB	31 Jan 2018 12:55:09	U	SMPL
J1364-17MSD - 1	MW-07MSD	16110	4.4521	-	7470A	PPB	31 Jan 2018 12:57:18	U	SMPL
J1365-02 - 1	GROUNDWATER	-279	-0.1043	-	7470A	PPB	31 Jan 2018 12:59:25	U	SMPL
CCV39 - 1	CCV39	18265	5.0512	-	7470A	PPB	31 Jan 2018 13:01:32	U	SMPL
CCB39 - 1	CCB39	-85	-0.0504	-	7470A	PPB	31 Jan 2018 13:03:39	U	SMPL
J1343-01 - 1	MH-6	-186	-0.0785	-	7470A	PPB	31 Jan 2018 13:05:47	U	SMPL
J1019-12LX5 - 1		-74	-0.0474	-	7470A	PPB	31 Jan 2018 13:07:54	U	SMPL
J1019-12A - 1		14114	3.8971	-	7470A	PPB	31 Jan 2018 13:10:01	U	SMPL
J1364-17LX5 - 1		-127	-0.0621	-	7470A	PPB	31 Jan 2018 13:12:09	U	SMPL
J1364-17A - 1		16909	4.6742	-	7470A	PPB	31 Jan 2018 13:14:16	U	SMPL
CCV40 - 1	CCV40	17686	4.8902	-	7470A	PPB	31 Jan 2018 13:16:24	U	SMPL
CCB40 - 1	CCB40	-101	-0.0549	-	7470A	PPB	31 Jan 2018 13:18:32	U	SMPL