

Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB93018

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP60984
calibration std. phosphate 1 ppm	WP61470
calibration std. phosphate 0.5 ppm	WP61469
calibration std. phosphate 0.3 ppm	WP61468
calibration std. 0 ppm	WP61465
phosphate ICV std.	WP61472
phosphate CCV std.	WP61471
calibration std. phosphate 0.1 ppm	WP61467
calibration std. phosphate 0.05 ppm	WP61466
Combined reagent	WP61476
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Sodium hydroxide, 1N	WP57036

Intercept: -0.0011

Slope: 0.6516

Regression: 0.999783

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	01/30/2018	11:49
2	CAL2	0.05	1	50	50	0.033	01/30/2018	11:49
3	CAL3	0.10	1	50	50	0.065	01/30/2018	11:50
4	CAL4	0.30	1	50	50	0.185	01/30/2018	11:50
5	CAL5	0.50	1	50	50	0.331	01/30/2018	11:51
6	CAL6	1.00	1	50	50	0.650	01/30/2018	11:51



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB93018

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.324	0.499	01/30/2018	11:52
2	ICB		1	50	50	0.000	0.002	01/30/2018	11:52
3	CCV1	0.50	1	50	50	0.326	0.502	01/30/2018	11:53
4	CCB1		1	50	50	0.000	0.002	01/30/2018	11:53
5	PB106229BL		1	50	50	0.000	0.002	01/30/2018	11:54
6	PB106229BS	0.50	1	50	50	0.316	0.487	01/30/2018	11:54
7	J1192-21		1	50	50	0.674	1.036	01/30/2018	11:55
8	J1347-01		1	50	50	1.226	1.883	01/30/2018	11:55
9	J1350-01		1	50	50	0.015	0.025	01/30/2018	11:56
10	J1350-02		1	50	50	0.016	0.026	01/30/2018	11:56
11	J1350-02DUP		1	50	50	0.016	0.026	01/30/2018	11:57
12	J1350-02MS	0.50	1	50	50	0.301	0.464	01/30/2018	11:57
13	J1350-02MSD	0.50	1	50	50	0.300	0.462	01/30/2018	11:58
14	J1192-21		2	50	50	0.398	0.612	01/30/2018	12:14
15	CCV2	0.50	1	50	50	0.326	0.502	01/30/2018	12:15
16	CCB2		1	50	50	0.000	0.002	01/30/2018	12:15
17	J1347-01		10	50	50	0.478	0.735	01/30/2018	12:16
18	CCV3	0.50	1	50	50	0.327	0.504	01/30/2018	12:17
19	CCB3		1	50	50	0.000	0.002	01/30/2018	12:17