

7471B


LB 930 23

INSTRUMENT ID: CVL

Linear

 μ Abs.:

42230



A= 0.0000e+000

B= 2.3668e-004

C= -5.3059e-002

Rho= 0.9997603

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.036	-0.036	71	0.000	71				
0.2	0.200	0.210	0.010	1110	0.0 %	1110	0			
2.5	2.500	2.577	0.077	11111	0.0 %	11111				
5.0	5.000	4.885	-0.115	20863	0.0 %	20863				
7.5	7.500	7.623	0.123	32434	0.0 %	32434				
10.0	10.000	9.942	-0.058	42230	0.0 %	42230				

LB93023 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 31 Jan 2018 15:08:05					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	71	-	0.0000	7471B	PPB	31 Jan 2018 15:11:55	S	Std
0.2 - 1	SO.2	1110	-	0.2000	7471B	PPB	31 Jan 2018 15:14:12	S	Std
2.5 - 1	SO.2	1111	-	2.5000	7471B	PPB	31 Jan 2018 15:16:30	S	Std
5.0 - 1	SO.5	20863	-	5.0000	7471B	PPB	31 Jan 2018 15:18:48	S	Std
7.5 - 1	SO.5	32434	-	7.5000	7471B	PPB	31 Jan 2018 15:21:04	S	Std
10.0 - 1	SO	42230	-	10.0000	7471B	PPB	31 Jan 2018 15:26:14	S	Std
ICV77 - 1	ICV77	16423	3.8339	-	7471B	PPB	31 Jan 2018 15:28:58	U	SMPL
ICB77 - 1	ICB77	-108	-0.0786	-	7471B	PPB	31 Jan 2018 15:31:14	U	SMPL
CCV41 - 1	CCV41	21063	4.9321	-	7471B	PPB	31 Jan 2018 15:33:29	U	SMPL
CCB41 - 1	CCB41	-86	-0.0734	-	7471B	PPB	31 Jan 2018 15:35:45	U	SMPL
CRA - 1	CRA	1111	0.2099	-	7471B	PPB	31 Jan 2018 15:38:01	U	SMPL
HighStd - 1	HighStd	43062	10.1388	-	7471B	PPB	31 Jan 2018 15:40:17	U	SMPL
ChkStd - 1	ChkStd	30078	7.0657	-	7471B	PPB	31 Jan 2018 15:42:33	U	SMPL
PB106206BL - 1	PBS	-80	-0.0720	-	7471B	PPB	31 Jan 2018 15:44:51	U	SMPL
PB106206BS - 1	LCSS	14870	3.4663	-	7471B	PPB	31 Jan 2018 15:47:08	U	SMPL
J1010-10 - 1	BU-02-013018	2917	0.6373	-	7471B	PPB	31 Jan 2018 15:49:25	U	SMPL
J1015-20 - 1	CL-02-013018-A	7433	1.7062	-	7471B	PPB	31 Jan 2018 15:51:42	U	SMPL
J1015-22 - 1	CL-02-013018-B	8358	1.9251	-	7471B	PPB	31 Jan 2018 15:53:59	U	SMPL
J1015-24 - 1	CL-02-013018-C	8378	1.9298	-	7471B	PPB	31 Jan 2018 15:56:17	U	SMPL
J1015-26 - 1	CL-02-013018-D	5856	1.3329	-	7471B	PPB	31 Jan 2018 15:58:33	U	SMPL
CCV42 - 1	CCV42	21009	4.9193	-	7471B	PPB	31 Jan 2018 16:00:49	U	SMPL
CCB42 - 1	CCB42	-84	-0.0729	-	7471B	PPB	31 Jan 2018 16:03:05	U	SMPL
J1016-07 - 1	OK-03-013018	939	0.1692	-	7471B	PPB	31 Jan 2018 16:05:21	U	SMPL
J1019-11 - 1	EO-03-012918-A	3934	0.8780	-	7471B	PPB	31 Jan 2018 16:07:37	U	SMPL
J1019-13 - 1	EO-03-012918-B	1758	0.3630	-	7471B	PPB	31 Jan 2018 16:09:53	U	SMPL
J1363-01 - 1	RB-01	41562	9.7837	-	7471B	PPB	31 Jan 2018 16:12:09	U	SMPL
J1367-01 - 1	EPE-106-5(100-102)	754	0.1254	-	7471B	PPB	31 Jan 2018 16:17:00	U	SMPL
J1367-01DUP - 1	EPE-106-5(100-102)DUP	769	0.1289	-	7471B	PPB	31 Jan 2018 16:19:17	U	SMPL
J1367-01MS - 1	EPE-106-5(100-102)MS	16224	3.7868	-	7471B	PPB	31 Jan 2018 16:21:35	U	SMPL
J1367-01MSD - 1	EPE-106-5(100-102)MSD	16719	3.9040	-	7471B	PPB	31 Jan 2018 16:23:52	U	SMPL
J1377-01 - 1	STONE	-14	-0.0564	-	7471B	PPB	31 Jan 2018 16:26:10	U	SMPL
J1378-01 - 1	RO-72	339	0.0272	-	7471B	PPB	31 Jan 2018 16:28:26	U	SMPL
CCV43 - 1	CCV43	21184	4.9607	-	7471B	PPB	31 Jan 2018 16:30:42	U	SMPL
CCB43 - 1	CCB43	-65	-0.0684	-	7471B	PPB	31 Jan 2018 16:32:58	U	SMPL
J1383-01 - 1	VNJ-218	2131	0.4513	-	7471B	PPB	31 Jan 2018 16:35:14	U	SMPL
J1383-04 - 1	SP-A	1558	0.3157	-	7471B	PPB	31 Jan 2018 16:37:31	U	SMPL
J1384-02 - 1	OILY-STONES-1	163	-0.0145	-	7471B	PPB	31 Jan 2018 16:39:47	U	SMPL
J1384-08 - 1	OILY-STONES-2	134	-0.0213	-	7471B	PPB	31 Jan 2018 16:42:03	U	SMPL
J1384-14 - 1	OILY-CONCRETE	152	-0.0171	-	7471B	PPB	31 Jan 2018 16:44:20	U	SMPL
J1385-01 - 1	EPE-106-5(120-122)	142	-0.0195	-	7471B	PPB	31 Jan 2018 16:46:37	U	SMPL
J1367-01LX5 - 1		158	-0.0157	-	7471B	PPB	31 Jan 2018 16:48:54	U	SMPL
J1367-01A - 1		18052	4.2194	-	7471B	PPB	31 Jan 2018 16:51:11	U	SMPL
CCV44 - 1	CCV44	20854	4.8826	-	7471B	PPB	31 Jan 2018 16:53:29	U	SMPL
CCB44 - 1	CCB44	-57	-0.0665	-	7471B	PPB	31 Jan 2018 16:55:45	U	SMPL