



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB93025

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	01/31/2018	15:08
2	J1407-01		1	>100.0	>212.0	01/31/2018	15:32
3	J1407-01DUP		1	>100.0	>212.0	01/31/2018	15:50

Result = (Celsius * 1.8) + 32

LB93025

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP13118-1

WorkList ID : 108137

Date : 1/31/2018 2:40:33 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/01/2018	Water	J1407-01 F	Flash Point	Cool 4 deg C	ENTA04	J13	TW-01	01/31/2018	1010A

Date/Time 1/31/18 2:45 PM
Received by: KS
Relinquished by: CP

Date/Time 1/31/18 15:55
Received by: CP
Relinquished by: KS