



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93029

Temp.: 21.3 °C

BalanceID: WC SC-4

SLOPE.: 97.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	01/31/2018	16:45
2	CAL2	1	Water			7.01	01/31/2018	16:46
3	CAL3	1	Water			10.03	01/31/2018	16:47
4	ICV	1	Water			7.00	01/31/2018	16:48
5	CCV1	1	Water			2.02	01/31/2018	16:50
6	J1014-04	1	Solid	20.02	20	8.15	01/31/2018	16:51
7	J1014-04DUP	1	Solid	20.03	20	8.16	01/31/2018	17:00
8	J1022-13	1	Solid	20.05	20	7.54	01/31/2018	17:01
9	J1413-06	1	Solid	20.03	20	9.20	01/31/2018	17:02
10	J1413-11	1	Solid	20.05	20	9.24	01/31/2018	17:05
11	CCV2	1	Water			12.01	01/31/2018	17:06

LB 03024

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I1413 WorkList ID : 108146 Date : 1/31/2018 4:25:47 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/31/2018	Solid	J1014-04	Corrosivity	Cool 4 deg C	PSEG05		HD-02-013118	01/31/2018	9045C
01/31/2018	Solid	J1022-13	Corrosivity	Cool 4 deg C	PSEG05		SU-04-013118	01/31/2018	9045C
01/31/2018	Solid	J1413-06	Corrosivity	Cool 4 deg C	PSEG02	J61	SOIL-1	01/31/2018	9045C
01/31/2018	Solid	J1413-11	Corrosivity	Cool 4 deg C	PSEG02	J61	SOIL-2	01/31/2018	9045C

Date/Time 01-31-18 6:25 AM
Received by: JP
Relinquished by: JP

Date/Time 01-31-18 6:25 AM
Received by: JP
Relinquished by: JP