



Analytical Summary Report

Analysis Method: 7196A
Parameter: Hexavalent Chromium
Run Number: LB93035

ANALYST: Indiana
SUPERVISOR REVIEW BY: apatel
pH Meter ID: WC pH Meter-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include Calibration Std. hexchrome 0.1 ppm, 5N sulfuric acid, Hexavalent Chromium ICV Std., etc.

Intercept: -0.0001 Slope: 0.7666 Regression: 0.999978

Table with 12 columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml/gm), Final Vol (ml), pH HN03, pH H2SO4, Absorb. at 540 (Backgrnd, Color), Absorbance Difference, Anal Date, Anal Time. Contains 7 rows of calibration data.



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Indiana

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: apatel

Run Number: LB93035

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100		2.11	0.000	0.387	0.387	0.505	01/31/2018	12:33
2	ICB		1	100	100		1.69	0.000	0.000	0.000	0.000	01/31/2018	12:34
3	CCV1	0.5	1	100	100		2.45	0.000	0.387	0.387	0.505	01/31/2018	12:34
4	CCB1		1	100	100		1.77	0.000	0.000	0.000	0.000	01/31/2018	12:35
5	PB106233BL		1	2.52	100	7.10	1.63	0.000	0.000	0.000	0.000	01/31/2018	12:35
6	PB106233BS	20	1	2.53	100	7.55	2.28	0.000	0.379	0.379	0.495	01/31/2018	12:36
7	J1010-10		1	2.54	100	7.60	2.34	0.016	0.016	0.000	0.000	01/31/2018	12:36
8	J1015-20		1	2.51	100	7.23	1.69	0.019	0.019	0.000	0.000	01/31/2018	12:37
9	J1015-22		1	2.52	100	7.32	1.96	0.021	0.021	0.000	0.000	01/31/2018	12:37
10	J1015-24		1	2.53	100	7.41	2.22	0.014	0.014	0.000	0.000	01/31/2018	12:38
11	J1015-26		1	2.51	100	7.69	2.08	0.043	0.043	0.000	0.000	01/31/2018	12:38
12	J1016-07		1	2.50	100	7.41	1.77	0.025	0.025	0.000	0.000	01/31/2018	12:39
13	J1023-07		1	2.53	100	7.78	1.92	0.027	0.027	0.000	0.000	01/31/2018	12:39
14	J1269-01		1	2.51	100	7.90	2.08	0.140	0.141	0.001	0.001	01/31/2018	12:40
15	CCV2	0.5	1	100	100		2.33	0.000	0.388	0.388	0.506	01/31/2018	12:40
16	CCB2		1	100	100		1.67	0.000	0.000	0.000	0.000	01/31/2018	12:41
17	J1377-01		1	2.52	100	7.88	1.90	0.007	0.007	0.000	0.000	01/31/2018	12:41
18	J1378-01		1	2.51	100	7.27	1.81	0.088	0.088	0.000	0.000	01/31/2018	12:42
19	J1378-01DUP		1	2.54	100	7.33	2.06	0.089	0.089	0.000	0.000	01/31/2018	12:42
20	J1378-01MSPre	40	2	2.55	100	7.60	2.44	0.091	0.125	0.034	0.044	01/31/2018	12:43
21	J1378-01MS2Ins	1284	40	2.50	100	7.32	2.16	0.086	0.624	0.538	0.702	01/31/2018	12:43
22	J1378-01MS3Post	40	2	2.51	100	7.44	1.62	0.095	0.201	0.106	0.138	01/31/2018	12:44
23	J1383-01		1	2.53	100	7.48	1.58	0.074	0.074	0.000	0.000	01/31/2018	12:44
24	J1383-04		1	2.54	100	7.63	2.02	0.090	0.090	0.000	0.000	01/31/2018	12:45
25	J1384-02		1	2.50	100	7.10	2.34	0.076	0.076	0.000	0.000	01/31/2018	12:45
26	J1384-08		1	2.51	100	7.17	2.41	0.083	0.083	0.000	0.000	01/31/2018	12:46
27	CCV3	0.5	1	100	100		2.47	0.000	0.388	0.388	0.506	01/31/2018	12:46
28	CCB3		1	100	100		1.56	0.000	0.000	0.000	0.000	01/31/2018	12:47
29	J1384-14		1	2.55	100	7.21	1.69	0.095	0.095	0.000	0.000	01/31/2018	12:47
30	CCV4	0.5	1	100	100		1.77	0.000	0.388	0.388	0.506	01/31/2018	12:48
31	CCB4		1	100	100		2.06	0.000	0.000	0.000	0.000	01/31/2018	12:48