



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93042

Temp.: 24.6 °C

BalanceID: WC SC-4

SLOPE.: 98.1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2270         |
| buffer solution pH 7 yellow    | W2159         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2220         |
| buffer solution pH 7 yellow    | W2219         |
| pH 2.00 Buffer                 | W2289         |
| Buffer Solution, PH12 (500ml)  | W2267         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

| Seq | LabID    | DF | Matrix | Weight (gm) | Volume (ml) | Result (pH) | Anal Date  | Anal Time |
|-----|----------|----|--------|-------------|-------------|-------------|------------|-----------|
| 1   | CAL1     | 1  | Water  |             |             | 4.01        | 01/29/2018 | 18:00     |
| 2   | CAL2     | 1  | Water  |             |             | 7.01        | 01/29/2018 | 18:01     |
| 3   | CAL3     | 1  | Water  |             |             | 10.03       | 01/29/2018 | 18:02     |
| 4   | ICV      | 1  | Water  |             |             | 6.99        | 01/29/2018 | 18:05     |
| 5   | CCV1     | 1  | Water  |             |             | 2.01        | 01/29/2018 | 18:06     |
| 6   | J1192-06 | 1  | Solid  | 20.03       | 20          | 10.28       | 01/29/2018 | 18:12     |
| 7   | CCV2     | 1  | Water  |             |             | 12.01       | 01/29/2018 | 18:25     |