



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB93052

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	02/01/2018	12:32
2	J1192-03		1	58.0	136.4	02/01/2018	13:05

Result = (Celsius * 1.8) + 32