


7471B

LB 93054
INSTRUMENT ID: CVL

Linear

 μ Abs.:

44026



A= 0.0000e+000

B= 2.2719e-004

C= -1.7307e-002

Rho= 0.9999891

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.025	-0.025	-34	0.000	-34				
0.2	0.200	0.219	0.019	1038	0.0 %	1038	0			
2.5	2.500	2.507	0.007	11112	0.0 %	11112				
5.0	5.000	4.992	-0.008	22050	0.0 %	22050				
7.5	7.500	7.522	0.022	33184	0.0 %	33184				
10.0	10.000	9.985	-0.015	44026	0.0 %	44026				

LB93054 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Feb 2018 15:29:19

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-34	-	0.00007471B	PPB		01 Feb 2018 15:31:48	S	Std
0.2 - 1	50.2	1038	-	0.20007471B	PPB		01 Feb 2018 15:34:05	S	Std
2.5 - 1	52.5	11112	-	2.50007471B	PPB		01 Feb 2018 15:36:22	S	Std
5.0 - 1	55	22050	-	5.00007471B	PPB		01 Feb 2018 15:38:40	S	Std
7.5 - 1	57.5	33184	-	7.50007471B	PPB		01 Feb 2018 15:40:57	S	Std
10.0 - 1	59	44026	-	10.00007471B	PPB		01 Feb 2018 15:45:52	S	Std
ICV79 - 1	ICV79	17245	3.9006	-7471B	PPB		01 Feb 2018 15:48:35	U	SMPL
ICB79 - 1	ICB79	-129	-0.0466	-7471B	PPB		01 Feb 2018 15:50:51	U	SMPL
CCV50 - 1	CCV50	22456	5.0845	-7471B	PPB		01 Feb 2018 15:53:06	U	SMPL
CCB50 - 1	CCB50	-129	-0.0466	-7471B	PPB		01 Feb 2018 15:55:22	U	SMPL
CRA - 1	CRA	1044	0.2199	-7471B	PPB		01 Feb 2018 15:57:38	U	SMPL
HighStd - 1	HighStd	44488	10.0901	-7471B	PPB		01 Feb 2018 15:59:54	U	SMPL
ChkStd - 1	ChkStd	31691	7.1827	-7471B	PPB		01 Feb 2018 16:02:10	U	SMPL
PB106239BL - 1	PBS	-199	-0.0625	-7471B	PPB		01 Feb 2018 16:04:28	U	SMPL
PB106239BS - 1	LCSS	15141	3.4226	-7471B	PPB		01 Feb 2018 16:06:45	U	SMPL
J1014-03 - 1	HD-02-013118	5114	1.1446	-7471B	PPB		01 Feb 2018 16:09:03	U	SMPL
J1022-12 - 1	SU-04-013118	3343	0.7422	-7471B	PPB		01 Feb 2018 16:11:19	U	SMPL
J1023-07 - 1	TR-04-013018	1235	0.2633	-7471B	PPB		01 Feb 2018 16:13:37	U	SMPL
J1405-01 - 1	TP-03(6-WATERLINE-BOTTOM@6)	2094	0.4584	-7471B	PPB		01 Feb 2018 16:15:55	U	SMPL
J1405-02 - 1	TP-04(CB#3-LINE-BOTTOM@4)	1558	0.3367	-7471B	PPB		01 Feb 2018 16:18:10	U	SMPL
CCV51 - 1	CCV51	22287	5.0461	-7471B	PPB		01 Feb 2018 16:20:27	U	SMPL
CCB51 - 1	CCB51	-130	-0.0466	-7471B	PPB		01 Feb 2018 16:22:43	U	SMPL
J1405-03 - 1	TP-05(CB#1-LINE@5)	492	0.0945	-7471B	PPB		01 Feb 2018 16:24:59	U	SMPL
J1405-04 - 1	TP-06(12-ST-MAIN-N-OF-MH#2-@6.5)	1031	0.2169	-7471B	PPB		01 Feb 2018 16:27:15	U	SMPL
J1405-05 - 1	TP-07(12-ST-LATERAL-LINE-E-OF-MH#2-@6)	1231	0.2624	-7471B	PPB		01 Feb 2018 16:29:31	U	SMPL
J1405-06 - 1	TP-08(4-15-TRENCH-E-OF-ST-MH#3-@6)	1362	0.2921	-7471B	PPB		01 Feb 2018 16:31:48	U	SMPL
J1405-07 - 1	TP-09(8-ST-E-OF-CB#4-@3.6)	1617	0.3501	-7471B	PPB		01 Feb 2018 16:34:05	U	SMPL
J1408-01 - 1	EPE-106-5(135-137)	46	-0.0069	-7471B	PPB		01 Feb 2018 16:36:21	U	SMPL
J1408-01DUP - 1	EPE-106-5(135-137)DUP	59	-0.0039	-7471B	PPB		01 Feb 2018 16:38:39	U	SMPL
J1408-01MS - 1	EPE-106-5(135-137)MS	16518	3.7355	-7471B	PPB		01 Feb 2018 16:40:56	U	SMPL
J1408-01MSD - 1	EPE-106-5(135-137)MSD	16478	3.7264	-7471B	PPB		01 Feb 2018 16:43:14	U	SMPL
J1408-02 - 1	EPE-106-5(140-142)	-41	-0.0266	-7471B	PPB		01 Feb 2018 16:45:30	U	SMPL
CCV52 - 1	CCV52	22342	5.0586	-7471B	PPB		01 Feb 2018 16:47:46	U	SMPL
CCB52 - 1	CCB52	-130	-0.0466	-7471B	PPB		01 Feb 2018 16:50:03	U	SMPL
J1410-02 - 1	RT-1495	1278	0.2730	-7471B	PPB		01 Feb 2018 16:52:20	U	SMPL
J1413-04 - 1	SOIL-1	1337	0.2865	-7471B	PPB		01 Feb 2018 16:54:36	U	SMPL
J1413-09 - 1	SOIL-2	1696	0.3680	-7471B	PPB		01 Feb 2018 16:56:52	U	SMPL
J1408-01LX5 - 1		15	-0.0139	-7471B	PPB		01 Feb 2018 17:04:55	U	SMPL
J1408-01A - 1		17320	3.9177	-7471B	PPB		01 Feb 2018 17:07:11	U	SMPL
CCV53 - 1	CCV53	22169	5.0193	-7471B	PPB		01 Feb 2018 17:09:28	U	SMPL
CCB53 - 1	CCB53	-121	-0.0448	-7471B	PPB		01 Feb 2018 17:11:46	U	SMPL