



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93063

Temp.: 18.3 °C

SLOPE.: 94.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/02/2018	10:20
2	CAL2	1	Water			7.00	02/02/2018	10:22
3	CAL3	1	Water			10.01	02/02/2018	10:24
4	ICV	1	Water			7.00	02/02/2018	10:26
5	CCV1	1	Water			2.00	02/02/2018	10:28
6	J1407-01	1	Water			7.60	02/02/2018	10:30
7	J1407-01DUP	1	Water			7.61	02/02/2018	10:32
8	CCV2	1	Water			12.00	02/02/2018	10:35

WORKLIST(Hardcopy Internal Chain)

LB93063

WorkList Name : PHJ1407

WorkList ID : 108205

Date : 2/2/2018 10:13:57 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/31/2018	Water	J1407-01	pH	Cool 4 deg C	ENTA04		TW-01	01/31/2018	SM 4500-PH B

Date/Time

02/02/18 10:15 AM

Received by:

SP

Relinquished by:

SP

Date/Time

02/02/18 10:45 AM

Received by:

SP

Relinquished by:

SP