


245.1

LB93065
INSTRUMENT ID: CV4

Linear

 μ Abs.:

35494



A= 0.0000e+000

B= 2.8341e-004

C= -2.5118e-002

Rho= 0.9999691

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	35	0.000	35				
0.05	0.050	0.051	0.001	268	0.0 %	268				
0.2	0.200	0.215	0.015	847	0.0 %	847				
2.5	2.500	2.508	0.008	8939	0.0 %	8939				
5.0	5.000	5.020	0.020	17800	0.0 %	17800				
7.5	7.500	7.438	-0.062	26332	0.0 %	26332				
10.0	10.000	10.034	0.034	35494	0.0 %	35494				

LB93065 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Feb 2018 12:37:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-146	-	0.0000	7470A	PPB	02 Feb 2018 13:14:04	S	Std
0.2 - 1	50.05	893	-	0.2000	7470A	PPB	02 Feb 2018 13:16:12	S	Std
2.5 - 1	50.2	9277	-	2.5000	7470A	PPB	02 Feb 2018 13:18:21	S	Std
5.0 - 1	52.5	17896	-	5.0000	7470A	PPB	02 Feb 2018 13:20:29	S	Std
7.5 - 1	55	27816	-	7.5000	7470A	PPB	02 Feb 2018 13:22:37	S	Std
10.0 - 1	57.5	35989	-	10.0000	7470A	PPB	02 Feb 2018 13:24:44	S	Std
ICV80 - 1	ICV80	13822	3.7959	-7470A	PPB	02 Feb 2018 13:27:29	U	SMPL	
ICB80 - 1	ICB80	-282	-0.0934	-7470A	PPB	02 Feb 2018 13:29:34	U	SMPL	
CCV54 - 1	CCV54	18902	5.1967	-7470A	PPB	02 Feb 2018 13:31:41	U	SMPL	
CCB54 - 1	CCB54	-278	-0.0923	-7470A	PPB	02 Feb 2018 13:33:48	U	SMPL	
CRA - 1	CRA	862	0.2221	-7470A	PPB	02 Feb 2018 13:35:54	U	SMPL	
HighStd - 1	HighStd	36167	9.9576	-7470A	PPB	02 Feb 2018 13:38:01	U	SMPL	
ChkStd - 1	ChkStd	26687	7.3435	-7470A	PPB	02 Feb 2018 13:40:07	U	SMPL	
PB106262BL - 1	PBW	-181	-0.0655	-7470A	PPB	02 Feb 2018 13:42:16	U	SMPL	
PB106262BS - 1	LCSW	12551	3.4454	-7470A	PPB	02 Feb 2018 13:44:24	U	SMPL	
PB106262TB - 1	PB106235TB	-133	-0.0523	-7470A	PPB	02 Feb 2018 13:46:32	U	SMPL	
J1014-04 - 1	HD-02-013118	-74	-0.0360	-7470A	PPB	02 Feb 2018 13:48:40	U	SMPL	
J1014-04DUP - 1	HD-02-013118DUP	-64	-0.0332	-7470A	PPB	02 Feb 2018 13:50:48	U	SMPL	
J1014-04MS - 1	HD-02-013118MS	13241	3.6357	-7470A	PPB	02 Feb 2018 13:52:57	U	SMPL	
J1014-04MSD - 1	HD-02-013118MSD	13652	3.7490	-7470A	PPB	02 Feb 2018 13:55:04	U	SMPL	
CCV55 - 1	CCV55	18774	5.1614	-7470A	PPB	02 Feb 2018 13:57:11	U	SMPL	
CCB55 - 1	CCB55	-287	-0.0947	-7470A	PPB	02 Feb 2018 13:59:17	U	SMPL	
J1022-13 - 1	SU-04-013118	-114	-0.0470	-7470A	PPB	02 Feb 2018 14:01:25	U	SMPL	
J1410-01 - 1	RT-1495	-58	-0.0316	-7470A	PPB	02 Feb 2018 14:03:32	U	SMPL	
J1410-04 - 1	72-11935	-50	-0.0294	-7470A	PPB	02 Feb 2018 14:05:39	U	SMPL	
J1410-05 - 1	CHRT-27166	-73	-0.0357	-7470A	PPB	02 Feb 2018 14:07:46	U	SMPL	
J1413-06 - 1	SOIL-1	-59	-0.0319	-7470A	PPB	02 Feb 2018 14:09:53	U	SMPL	
J1413-11 - 1	SOIL-2	-72	-0.0354	-7470A	PPB	02 Feb 2018 14:12:01	U	SMPL	
PB106263BL - 1	PBW	-59	-0.0319	-7470A	PPB	02 Feb 2018 14:14:08	U	SMPL	
PB106263BS - 1	LCSW	13328	3.6597	-7470A	PPB	02 Feb 2018 14:16:17	U	SMPL	
PB106263TB - 1	PB106237TB	-282	-0.0934	-7470A	PPB	02 Feb 2018 14:18:25	U	SMPL	
J1414-01 - 1	SB1R-0.5-1.0	1535	0.4077	-7470A	PPB	02 Feb 2018 14:20:32	U	SMPL	
CCV56 - 1	CCV56	18546	5.0985	-7470A	PPB	02 Feb 2018 14:22:39	U	SMPL	
CCB56 - 1	CCB56	-314	-0.1022	-7470A	PPB	02 Feb 2018 14:24:45	U	SMPL	
J1414-02 - 1	SB2R-0.5-1.0	117	0.0167	-7470A	PPB	02 Feb 2018 14:26:53	U	SMPL	
J1414-03 - 1	SB9-0.0-0.5	36	-0.0057	-7470A	PPB	02 Feb 2018 14:29:39	U	SMPL	
J1414-03DUP - 1	SB9-0.0-0.5DUP	28	-0.0079	-7470A	PPB	02 Feb 2018 14:31:45	U	SMPL	
J1414-03MS - 1	SB9-0.0-0.5MS	15490	4.2558	-7470A	PPB	02 Feb 2018 14:33:52	U	SMPL	
J1414-03MSD - 1	SB9-0.0-0.5MSD	15946	4.3816	-7470A	PPB	02 Feb 2018 14:35:59	U	SMPL	
J1014-04LX5 - 1		-226	-0.0779	-7470A	PPB	02 Feb 2018 14:38:07	U	SMPL	
J1014-04A - 1		14217	3.9048	-7470A	PPB	02 Feb 2018 14:40:15	U	SMPL	
J1414-03LX5 - 1		-222	-0.0768	-7470A	PPB	02 Feb 2018 14:42:23	U	SMPL	
J1414-03A - 1		15390	4.2283	-7470A	PPB	02 Feb 2018 14:44:31	U	SMPL	
CCV57 - 1	CCV57	18500	5.0859	-7470A	PPB	02 Feb 2018 14:46:38	U	SMPL	
CCB57 - 1	CCB57	-270	-0.0900	-7470A	PPB	02 Feb 2018 14:48:45	U	SMPL	