



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93066

Temp.: 20.7 °C

SLOPE.: 94.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/02/2018	13:15
2	CAL2	1	Water			7.00	02/02/2018	13:17
3	CAL3	1	Water			10.01	02/02/2018	13:19
4	ICV	1	Water			7.00	02/02/2018	13:21
5	CCV1	1	Water			2.00	02/02/2018	13:23
6	J1439-01	1	Water			5.52	02/02/2018	13:25
7	J1439-01DUP	1	Water			5.53	02/02/2018	13:27
8	CCV2	1	Water			12.00	02/02/2018	13:30

WORKLIST(Hardcopy Internal Chain)

LB93066

WorkList Name : pHJ1439

WorkList ID : 108214

Date : 2/2/2018 1:07:47 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/29/2018	Water	J1439-01	pH	Cool 4 deg C	STER13	O33	STERIGENICS-TUSTIN-WATT	01/29/2018	SM 4500-PH B

Date/Time

02/02/18 1:10 PM

Received by:

SP

Relinquished by:

SP

Date/Time

02/02/18 1:40 PM

Received by:

SP

Relinquished by:

SP