


245.1

LB 93079
(INSTRUMENT ID: CV)

Linear

 μ Abs.:

35494



A= 0.0000e+000

B= 2.8341e-004

C= -2.5118e-002

Rho= 0.9999691

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	35	0.000	35				
0.05	0.050	0.051	0.001	268	0.0 %	268				
0.2	0.200	0.215	0.015	847	0.0 %	847				
2.5	2.500	2.508	0.008	8939	0.0 %	8939				
5.0	5.000	5.020	0.020	17800	0.0 %	17800				
7.5	7.500	7.438	-0.062	26332	0.0 %	26332				
10.0	10.000	10.034	0.034	35494	0.0 %	35494				

LB93079 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 02 Feb 2018 15:04:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	35	-	0.0000	245.1	PPB	02 Feb 2018 15:07:09	S	Std
0.05 - 1	50.05	268	-	0.0500	245.1	PPB	02 Feb 2018 15:09:17	S	Std
0.2 - 1	50.2	847	-	0.2000	245.1	PPB	02 Feb 2018 15:11:25	S	Std
2.5 - 1	52.5	8939	-	2.5000	245.1	PPB	02 Feb 2018 15:13:33	S	Std
5.0 - 1	55	17800	-	5.0000	245.1	PPB	02 Feb 2018 15:15:41	S	Std
7.5 - 1	57.5	26332	-	7.5000	245.1	PPB	02 Feb 2018 15:17:49	S	Std
10.0 - 1	510	35494	-	10.0000	245.1	PPB	02 Feb 2018 15:19:56	S	Std
ICV81 - 1	ICV81	13577	3.8227	-245.1		PPB	02 Feb 2018 15:22:37	U	SMPL
ICB81 - 1	ICB81	-153	-0.0685	-245.1		PPB	02 Feb 2018 15:24:43	U	SMPL
CCV58 - 1	CCV58	17583	4.9580	-245.1		PPB	02 Feb 2018 15:26:50	U	SMPL
CCB58 - 1	CCB58	-167	-0.0724	-245.1		PPB	02 Feb 2018 15:28:56	U	SMPL
CRA - 1	CRA	828	0.2095	-245.1		PPB	02 Feb 2018 15:31:04	U	SMPL
HighStd - 1	HighStd	35893	10.1472	-245.1		PPB	02 Feb 2018 15:33:10	U	SMPL
ChkStd - 1	ChkStd	24428	6.8979	-245.1		PPB	02 Feb 2018 15:35:17	U	SMPL
PB106282BL - 1	PBW	-260	-0.0988	-245.1		PPB	02 Feb 2018 15:37:26	U	SMPL
PB106282BS - 1	LCSW	13381	3.7671	-245.1		PPB	02 Feb 2018 15:39:35	U	SMPL
J1435-01 - 1	COMP	-159	-0.0702	-245.1		PPB	02 Feb 2018 15:44:23	U	SMPL
J1435-01DUP - 1	COMPDUP	-146	-0.0665	-245.1		PPB	02 Feb 2018 15:46:30	U	SMPL
J1435-01MS - 1	COMPMS	13031	3.6680	-245.1		PPB	02 Feb 2018 15:48:38	U	SMPL
J1435-01MSD - 1	COMPMSD	12848	3.6161	-245.1		PPB	02 Feb 2018 15:50:46	U	SMPL
J1262-01 - 1	BELL-0118	-354	-0.1254	-245.1		PPB	02 Feb 2018 15:52:53	U	SMPL
CCV59 - 1	CCV59	17883	5.0430	-245.1		PPB	02 Feb 2018 15:54:59	U	SMPL
CCB59 - 1	CCB59	-178	-0.0756	-245.1		PPB	02 Feb 2018 15:57:06	U	SMPL
J1264-01 - 1	2-AVE-09	-194	-0.0801	-245.1		PPB	02 Feb 2018 15:59:14	U	SMPL
J1435-01LX5 - 1		-65	-0.0435	-245.1		PPB	02 Feb 2018 16:03:57	U	SMPL
J1435-01A - 1		13708	3.8598	-245.1		PPB	02 Feb 2018 16:06:04	U	SMPL
CCV60 - 1	CCV60	17666	4.9815	-245.1		PPB	02 Feb 2018 16:08:11	U	SMPL
CCB60 - 1	CCB60	-167	-0.0724	-245.1		PPB	02 Feb 2018 16:10:18	U	SMPL