



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93082

Temp.: 23.1 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/02/2018	16:40
2	CAL2	1	Water			6.99	02/02/2018	16:41
3	CAL3	1	Water			10.03	02/02/2018	16:42
4	ICV	1	Water			7.00	02/02/2018	16:43
5	CCV1	1	Water			2.01	02/02/2018	16:45
6	J1449-01	1	Solid	20.02	20	4.49	02/02/2018	16:46
7	J1449-02	1	Solid	20.04	20	4.75	02/02/2018	16:48
8	J1449-02DUP	1	Solid	20.05	20	4.76	02/02/2018	16:49
9	CCV2	1	Water			12.02	02/02/2018	16:50

1393082

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J1449

WorkList ID : 108220

Date : 2/2/2018 4:29:25 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/31/2018	Solid	J1449-01 ✓	pH	Cool 4 deg C	USAR03	Q23	HCLA-B2-1	01/31/2018	9045C
02/01/2018	Solid	J1449-02 ✓	pH	Cool 4 deg C	USAR03	Q23	HCLA-B5-1	02/01/2018	9045C

Date/Time 02/02/18 4:30 PM
Received by: SA
Relinquished by: CB

Date/Time 02/02/18 5:30 PM
Received by: CB
Relinquished by: JD