


ISM02.4

LB 93083
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45414



A= 0.0000e+000

B= 2.2109e-004

C= -4.2355e-002

Rho= 0.9999867

Accept=Accepted

slope
1.51e-5 per

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/d
0.00	0.000	-0.021	-0.021	96	0.000	96					—
0.2	0.200	0.202	0.002	1104	0.0 %	1104					i
2.5	2.500	2.538	0.038	11669	0.0 %	11669					2
5.0	5.000	4.982	-0.018	22727	0.0 %	22727					0
7.5	7.500	7.501	0.001	34121	0.0 %	34121					0
10.0	10.000	9.998	-0.002	45414	0.0 %	45414					0

103
2/2/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0.0	96	-		0 ISM02.4	PPB	2/2/2018 17:24	Std
	0.2 S0.2	1104	-		0.2 ISM02.4	PPB	2/2/2018 17:26	Std
	2.5 S2.5	11669	-		2.5 ISM02.4	PPB	2/2/2018 17:29	Std
	5 S5	22727	-		5 ISM02.4	PPB	2/2/2018 17:31	Std
	7.5 S7.5	34121	-		7.5 ISM02.4	PPB	2/2/2018 17:33	Std
	10 S10	45414	-		10 ISM02.4	PPB	2/2/2018 17:35	Std
ICV	ICV	17775	3.8875 -		ISM02.4	PPB	2/2/2018 17:38	SMPL
ICB	ICB	-11	-0.0448 -		ISM02.4	PPB	2/2/2018 17:40	SMPL
CCV061	CCV061	22706	4.9777 -		ISM02.4	PPB	2/2/2018 17:43	SMPL
CCB061	CCB061	-24	-0.0477 -		ISM02.4	PPB	2/2/2018 17:45	SMPL
PB106238BL	PBS004	140	-0.0114 -		ISM02.4	PPB	2/2/2018 17:47	SMPL
J1318-01	MF6C57	933	0.1639 -		ISM02.4	PPB	2/2/2018 17:49	SMPL
J1318-02	MF6C58	166	-0.0057 -		ISM02.4	PPB	2/2/2018 17:52	SMPL
J1318-03	MF6C59	173	-0.0041 -		ISM02.4	PPB	2/2/2018 17:54	SMPL
J1318-04	MF6C60	685	0.1091 -		ISM02.4	PPB	2/2/2018 17:56	SMPL
J1318-05	MF6C61	308	0.0257 -		ISM02.4	PPB	2/2/2018 17:59	SMPL
J1318-06	MF6C62	194	0.0005 -		ISM02.4	PPB	2/2/2018 18:01	SMPL
J1318-07	MF6C63	580	0.0859 -		ISM02.4	PPB	2/2/2018 18:03	SMPL
J1318-08	MF6C64	471	0.0618 -		ISM02.4	PPB	2/2/2018 18:05	SMPL
J1318-09	MF6C64D	340	0.0328 -		ISM02.4	PPB	2/2/2018 18:08	SMPL
J1318-10	MF6C64S	11609	2.5243 -		ISM02.4	PPB	2/2/2018 18:10	SMPL
CCV062	CCV062	22831	5.0053 -		ISM02.4	PPB	2/2/2018 18:19	SMPL
CCB062	CCB062	7	-0.0408 -		ISM02.4	PPB	2/2/2018 18:21	SMPL