



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93085

Temp.: 21.2 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 2.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	02/02/2018	17:35
2	CAL2	1	Water			7.01	02/02/2018	17:36
3	CAL3	1	Water			10.02	02/02/2018	17:37
4	ICV	1	Water			7.01	02/02/2018	17:38
5	CCV1	1	Water			2.02	02/02/2018	17:40
6	J1399-02	1	Solid	20.02	20	8.77	02/02/2018	17:41
7	J1399-02DUP	1	Solid	20.03	20	8.78	02/02/2018	17:42
8	J1399-04	1	Solid	20.05	20	7.99	02/02/2018	17:43
9	J1437-03	1	Solid	20.02	20	7.16	02/02/2018	17:46
10	J1442-01	1	Solid	20.03	20	8.49	02/02/2018	17:48
11	J1442-02	1	Solid	20.05	20	8.87	02/02/2018	17:50
12	J1442-03	1	Solid	20.01	20	8.21	02/02/2018	17:51
13	J1444-02	1	Solid	20.04	20	6.90	02/02/2018	17:52
14	J1445-03	1	Solid	20.02	20	7.13	02/02/2018	18:00
15	J1453-04	1	Solid	20.01	20	5.19	02/02/2018	18:01
16	CCV2	1	Water			12.02	02/02/2018	18:02
17	J1453-08	1	Solid	20.03	20	5.30	02/02/2018	18:03
18	J1453-12	1	Solid	20.05	20	5.53	02/02/2018	18:05
19	CCV3	1	Water			2.01	02/02/2018	18:06

1549085

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY J1445 WorkList ID : 108219 Date : 2/2/2018 2:41:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/02/2018	Solid	J1399-02	Corrosivity	Cool 4 deg C	PSEG05	J62	OR-01-020218-A	02/02/2018	9045C
02/02/2018	Solid	J1399-04	Corrosivity	Cool 4 deg C	PSEG05	J62	OR-01-020218-B	02/02/2018	9045C
02/02/2018	Solid	J1437-03	Corrosivity	Cool 4 deg C	PSEG01	J62	RO-148-20	02/02/2018	9045C
02/01/2018	Solid	J1442-01	Corrosivity	Cool 4 deg C	CLEA11	O33	SECOND-FLOOR-CHASSIS-A	02/01/2018	9045C
02/01/2018	Solid	J1442-02	Corrosivity	Cool 4 deg C	CLEA11	O33	FIRST-FLOOR-CHASSIS-WA	02/01/2018	9045C
02/01/2018	Solid	J1442-03	Corrosivity	Cool 4 deg C	CLEA11	O33	BUS-WASH-RECLAIM-SYST	02/01/2018	9045C
02/01/2018	Solid	J1444-02	Corrosivity	Cool 4 deg C	VIA501	O31	WC-PH3-MGP-03	02/01/2018	9045C
02/02/2018	Solid	J1445-03	Corrosivity	Cool 4 deg C	PSEG01	J62	M-4	02/02/2018	9045C
02/02/2018	Solid	J1453-04	Corrosivity	Cool 4 deg C	PSEG01	J62	P-1	02/02/2018	9045C
02/02/2018	Solid	J1453-08	Corrosivity	Cool 4 deg C	PSEG01	J62	P-2	02/02/2018	9045C
02/02/2018	Solid	J1453-12	Corrosivity	Cool 4 deg C	PSEG01	J62	P-3	02/02/2018	9045C

Date/Time 02/01/18 5:25 PM
Received by: JA
Relinquished by: CA

Date/Time 02/02/18 6:35 AM
Received by: CR
Relinquished by: JR