



Analytical Summary Report

Analysis Method: SM4500-H B

Analyst By : gorge

Parameter: Field pH

Supervisor Review By : apatel

Run Number: LB93088

Temp.: 5.50 °C

SLOPE.: 1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 07 Control Limits[+/- 0.1].

True Value of CCV2 = 07 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.02	01/26/2018	12:05
2	CAL2	1	Water			7.01	01/26/2018	12:07
3	CAL3	1	Water			10	01/26/2018	12:09
4	ICV	1	Water			7	01/26/2018	12:11
5	CCV1	1	Water			7.02	10/26/2018	12:19
6	J1350-01	1	Water			4.89	01/26/2018	12:30
7	J1350-02	1	Water			4.97	01/26/2018	12:40
8	CCV2	1	Water			6.98	01/26/2018	12:52