



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93104

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/05/2018	17:00
2	CAL2	1	Water			7.01	02/05/2018	17:01
3	CAL3	1	Water			10.03	02/05/2018	17:02
4	ICV	1	Water			7.00	02/05/2018	17:05
5	CCV1	1	Water			2.01	02/05/2018	17:06
6	J1459-04	1	Solid	20.02	20	8.01	02/05/2018	17:08
7	J1459-04DUP	1	Solid	20.03	20	8.02	02/05/2018	17:10
8	J1459-08	1	Solid	20.04	20	8.23	02/05/2018	17:11
9	J1460-04	1	Solid	20.01	20	7.01	02/05/2018	17:13
10	J1460-08	1	Solid	20.03	20	5.93	02/05/2018	17:15
11	CCV2	1	Water			12.02	02/05/2018	17:16

2393104

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY J1460 WorkList ID : 108258 Date : 2/5/2018 4:23:49 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/05/2018	Solid	J1459-04	Corrosivity	Cool 4 deg C	PSEG01	M53	TP-A	02/05/2018	9045C
02/05/2018	Solid	J1459-08	Corrosivity	Cool 4 deg C	PSEG01	M53	TP-B	02/05/2018	9045C
02/05/2018	Solid	J1460-04	Corrosivity	Cool 4 deg C	PSEG01	M51	B1	02/05/2018	9045C
02/05/2018	Solid	J1460-08	Corrosivity	Cool 4 deg C	PSEG01	M51	B2	02/05/2018	9045C

Date/Time 02-05-18 4:25pm
Received by: J1
Relinquished by: CP

Date/Time 02-05-18 5:35pm
Received by: CP
Relinquished by: J1