


7471B

LB 93107
INSTRUMENT ID: CVL

Linear

 μ Abs.:

42599



A= 0.0000e+000

B= 2.3466e-004

C= -5.1332e-002

Rho= 0.9998155

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.029	-0.029	94	0.000	94				
0.2	0.200	0.193	-0.007	1043	0.0 %	1043				
2.5	2.500	2.455	-0.045	10680	0.0 %	10680				
5.0	5.000	5.154	0.154	22184	0.0 %	22184				
7.5	7.500	7.482	-0.018	32101	0.0 %	32101				
10	10.000	9.945	-0.055	42599	0.0 %	42599				

LB93107 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Feb 2018 11:49:08

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	94	-	0.00007471B	PPB		06 Feb 2018 11:50:59	S	Std
0.2 - 1	50.2	1043	-	0.20007471B	PPB		06 Feb 2018 11:53:16	S	Std
2.5 - 1	52.5	10680	-	2.50007471B	PPB		06 Feb 2018 11:55:34	S	Std
5.0 - 1	55	22184	-	5.00007471B	PPB		06 Feb 2018 11:57:51	S	Std
7.5 - 1	57.5	32101	-	7.50007471B	PPB		06 Feb 2018 12:00:08	S	Std
10 - 1	510	42599	-	10.00007471B	PPB		06 Feb 2018 12:02:26	S	Std
ICV85 - 1	ICV85	16733	3.8753	-7471B	PPB		06 Feb 2018 12:05:55	U	SMPL
ICB85 - 1	ICB85	14	-0.0480	-7471B	PPB		06 Feb 2018 12:08:10	U	SMPL
CCV75 - 1	CCV75	20906	4.8545	-7471B	PPB		06 Feb 2018 12:10:26	U	SMPL
CCB75 - 1	CCB75	9	-0.0492	-7471B	PPB		06 Feb 2018 12:12:42	U	SMPL
CRA - 1	CRA	1039	0.1925	-7471B	PPB		06 Feb 2018 12:14:58	U	SMPL
HighStd - 1	HighStd	43021	10.0441	-7471B	PPB		06 Feb 2018 12:17:14	U	SMPL
ChkStd - 1	ChkStd	30222	7.0406	-7471B	PPB		06 Feb 2018 12:19:30	U	SMPL
PB106319BL - 1	PBS	52	-0.0391	-7471B	PPB		06 Feb 2018 12:24:26	U	SMPL
PB106319BS - 1	LCSS	15337	3.5477	-7471B	PPB		06 Feb 2018 12:26:42	U	SMPL
J1454-01 - 1	8-SAN-5-EW(0-4)	5001	1.1222	-7471B	PPB		06 Feb 2018 12:33:50	U	SMPL
J1454-01DUP - 1	8-SAN-5-EW(0-4)DUP	4534	1.0126	-7471B	PPB		06 Feb 2018 12:36:06	U	SMPL
J1454-02 - 1	J1454-01MS	22508	5.2304	-7471B	PPB		06 Feb 2018 12:38:23	U	SMPL
J1454-03 - 1	J1454-01MSD	21724	5.0465	-7471B	PPB		06 Feb 2018 12:40:41	U	SMPL
J1454-04 - 1	8-SAN-5-EW(4-8)	1295	0.2526	-7471B	PPB		06 Feb 2018 12:42:56	U	SMPL
CCV76 - 1	CCV76	20909	4.8552	-7471B	PPB		06 Feb 2018 12:45:13	U	SMPL
CCB76 - 1	CCB76	-4	-0.0523	-7471B	PPB		06 Feb 2018 12:47:29	U	SMPL
J1454-05 - 1	8-SAN-6-EW(2-4)	8142	1.8593	-7471B	PPB		06 Feb 2018 12:49:45	U	SMPL
J1454-06 - 1	8-SAN-6-EW(4-8)	4142	0.9206	-7471B	PPB		06 Feb 2018 12:52:01	U	SMPL
J1454-07 - 1	8-SAN-7-EW(2-4)	14015	3.2375	-7471B	PPB		06 Feb 2018 12:54:17	U	SMPL
J1454-08 - 1	8-SAN-7-EW(4-7)	2296	0.4875	-7471B	PPB		06 Feb 2018 12:56:34	U	SMPL
J1454-09 - 1	8-SAN-8-EW(1-4)	11538	2.6562	-7471B	PPB		06 Feb 2018 12:58:51	U	SMPL
J1454-10 - 1	8-SAN-8-EW(4-7)	12382	2.8543	-7471B	PPB		06 Feb 2018 13:01:08	U	SMPL
J1454-11 - 1	8-SAN-9-EW(1-4)	751	0.1249	-7471B	PPB		06 Feb 2018 13:03:25	U	SMPL
J1454-12 - 1	8-SAN-9-EW(4-6)	64920	15.1829	-7471B	PPB		06 Feb 2018 13:05:43	U	SMPL
J1454-13 - 1	8-SAN-9-WW(4-6)	28037	6.5279	-7471B	PPB		06 Feb 2018 13:08:01	U	SMPL
J1454-14 - 1	8-SAN-10-WW(1-4)	6033	1.3644	-7471B	PPB		06 Feb 2018 13:10:20	U	SMPL
CCV77 - 1	CCV77	20826	4.8357	-7471B	PPB		06 Feb 2018 13:12:37	U	SMPL
CCB77 - 1	CCB77	63	-0.0365	-7471B	PPB		06 Feb 2018 13:14:53	U	SMPL
J1454-15 - 1	8-SAN-10-WW(4-8)	2417	0.5158	-7471B	PPB		06 Feb 2018 13:17:10	U	SMPL
J1454-16 - 1	8-SAN-10-EW(2-4)	9583	2.1974	-7471B	PPB		06 Feb 2018 13:19:26	U	SMPL
J1454-17 - 1	8-SAN-10-EW(4-8)	11553	2.6597	-7471B	PPB		06 Feb 2018 13:21:42	U	SMPL
J1454-18 - 1	SA-ST-LATERAL-LINE-1-NW@3.5	6850	1.5561	-7471B	PPB		06 Feb 2018 13:23:59	U	SMPL
J1454-19 - 1	SA-ST-LATERAL-LINE-1-SW@3.5	2907	0.6308	-7471B	PPB		06 Feb 2018 13:26:16	U	SMPL
J1454-20 - 1	TP-03(6-WATER-LINE-BOTTEM@10)	1706	0.3490	-7471B	PPB		06 Feb 2018 13:28:32	U	SMPL
PB106321BL - 1	PBS	177	-0.0098	-7471B	PPB		06 Feb 2018 13:30:50	U	SMPL
PB106321BS - 1	LCSS	14579	3.3698	-7471B	PPB		06 Feb 2018 13:33:07	U	SMPL
J1456-01 - 1	8-SAN-9-WW(1-4)	7940	1.8119	-7471B	PPB		06 Feb 2018 13:35:25	U	SMPL
J1456-01DUP - 1	8-SAN-9-WW(1-4)DUP	7033	1.5990	-7471B	PPB		06 Feb 2018 13:37:41	U	SMPL
CCV78 - 1	CCV78	20851	4.8416	-7471B	PPB		06 Feb 2018 13:39:57	U	SMPL
CCB78 - 1	CCB78	57	-0.0380	-7471B	PPB		06 Feb 2018 13:42:13	U	SMPL
J1456-02 - 1	J1456-01MS	22434	5.2131	-7471B	PPB		06 Feb 2018 13:44:30	U	SMPL
J1456-03 - 1	J1456-01MSD	25188	5.8593	-7471B	PPB		06 Feb 2018 13:46:46	U	SMPL
J1456-04 - 1	8-SAN-6-EW(12-14)	10528	2.4192	-7471B	PPB		06 Feb 2018 13:49:03	U	SMPL
J1456-05 - 1	SA-ST-LATERAL-LINE-2-SW@4	1462	0.2917	-7471B	PPB		06 Feb 2018 13:51:20	U	SMPL
J1456-06 - 1	SA-ST-LATERAL-LINE-2-NW@4	1641	0.3337	-7471B	PPB		06 Feb 2018 13:53:37	U	SMPL
J1457-01 - 1	BROOK-101	18689	4.3343	-7471B	PPB		06 Feb 2018 13:55:53	U	SMPL
J1459-01 - 1	TP-A	1585	0.3206	-7471B	PPB		06 Feb 2018 13:58:11	U	SMPL
J1459-05 - 1	TP-B	3180	0.6949	-7471B	PPB		06 Feb 2018 14:00:28	U	SMPL
J1460-01 - 1	B1	326	0.0252	-7471B	PPB		06 Feb 2018 14:02:46	U	SMPL
J1460-05 - 1	B2	839	0.1455	-7471B	PPB		06 Feb 2018 14:05:03	U	SMPL
CCV79 - 1	CCV79	22183	5.1542	-7471B	PPB		06 Feb 2018 14:07:19	U	SMPL
CCB79 - 1	CCB79	77	-0.0333	-7471B	PPB		06 Feb 2018 14:09:35	U	SMPL
J1454-01LX5 - 1		1040	0.1927	-7471B	PPB		06 Feb 2018 14:11:52	U	SMPL
J1454-01A - 1		22064	5.1262	-7471B	PPB		06 Feb 2018 14:14:08	U	SMPL
J1456-01LX5 - 1		1415	0.2807	-7471B	PPB		06 Feb 2018 14:16:25	U	SMPL
J1456-01A - 1		23760	5.5242	-7471B	PPB		06 Feb 2018 14:18:41	U	SMPL
J1454-12DLX5 - 1		12696	2.9279	-7471B	PPB		06 Feb 2018 14:20:58	U	SMPL
CCV80 - 1	CCV80	22064	5.1262	-7471B	PPB		06 Feb 2018 14:23:16	U	SMPL
CCB80 - 1	CCB80	58	-0.0377	-7471B	PPB		06 Feb 2018 14:25:33	U	SMPL