

LB93109

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/6/2018 12:25

Test: ^{Reactive} Total CN Mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	99.986	0.0	0.091	
ICB1	0.285	0.0	0.004	
CCV1	252.141	0.0	0.224	
CCB1	0.122	0.0	0.004	
PB106334BL	0.094	0.0	0.004	
J1459-04	-0.060	0.0	0.003	
J1459-04DUP	-0.115	0.0	0.003	
J1459-08	0.160	0.0	0.004	
J1460-04	0.132	0.0	0.004	
J1460-08	0.287	0.0	0.004	
CCV2	255.046	0.0	0.226	
CCB2	0.392	0.0	0.004	

N	12
Mean	50.706
SD	98.9803
CV%	195.21

AS

Aquakem v. 7.2AQ1

Results from time period:

Tue Feb 06 08:39:52 2018

Tue Feb 06 12:16:39 2018

Sample Id	Sam/Ctr/c	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.0017	µg/l	2/6/2018 10:21:43	
5.OPPBCN	A	Reactive CN	P	5.2277	µg/l	2/6/2018 10:21:44	
10PPBCN	A	Reactive CN	P	10.165	µg/l	2/6/2018 10:21:45	
50PPBCN	A	Reactive CN	P	49.4106	µg/l	2/6/2018 10:21:46	
100PPBCN	A	Reactive CN	P	100.1794	µg/l	2/6/2018 10:21:47	
250PPBCN	A	Reactive CN	P	250.003	µg/l	2/6/2018 10:21:48	
500PPBCN	A	Reactive CN	P	500.016	µg/l	2/6/2018 10:21:49	
ICV1	S	Reactive CN	P	99.9862	µg/l	2/6/2018 12:09:40	
ICB1	S	Reactive CN	P	0.2852	µg/l	2/6/2018 12:09:41	
CCV1	S	Reactive CN	P	252.1405	µg/l	2/6/2018 12:09:43	
CCB1	S	Reactive CN	P	0.1219	µg/l	2/6/2018 12:09:45	
PB106334BL	S	Reactive CN	P	0.094	µg/l	2/6/2018 12:09:47	
J1459-04	S	Reactive CN	P	-0.0604	µg/l	2/6/2018 12:09:48	
J1459-04DUP	S	Reactive CN	P	-0.1147	µg/l	2/6/2018 12:09:49	
J1459-08	S	Reactive CN	P	0.1599	µg/l	2/6/2018 12:14:26	
J1460-04	S	Reactive CN	P	0.1316	µg/l	2/6/2018 12:14:27	
J1460-08	S	Reactive CN	P	0.2866	µg/l	2/6/2018 12:14:28	
CCV2	S	Reactive CN	P	255.0461	µg/l	2/6/2018 12:14:29	
CCB2	S	Reactive CN	P	0.392	µg/l	2/6/2018 12:14:30	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/6/2018 10:22


Test Total CN

Accepted 2/6/2018 10:22

Factor 1144
Bias 0.004

Coeff. of det. 0.999998

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.0017	0.0000	
2	5.0PPBCN	0.008	5.2277	5.0000	
3	10PPBCN	0.012	10.1650	10.0000	
4	50PPBCN	0.047	49.4106	50.0000	
5	100PPBCN	0.091	100.1794	100.0000	
6	250PPBCN	0.222	250.0030	250.0000	
7	500PPBCN	0.440	500.0160	500.0000	