



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93119

Temp.: 14.2 °C

SLOPE.: 97.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/06/2018	15:10
2	CAL2	1	Water			7.00	02/06/2018	15:11
3	CAL3	1	Water			10.02	02/06/2018	15:12
4	ICV	1	Water			6.99	02/06/2018	15:15
5	CCV1	1	Water			2.02	02/06/2018	15:16
6	J1466-02	1	Water			11.86	02/06/2018	15:18
7	J1466-02DUP	1	Water			11.87	02/06/2018	15:19
8	CCV2	1	Water			12.01	02/06/2018	15:20

LB93119

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY J1466 WorkList ID : 108290 Date : 2/6/2018 2:53:14 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/06/2018	Water	J1466-02 -C	Corrosivity	Cool 4 deg C	PSEG01	J63	FRAC-TANK	02/06/2018	9040C

Date/Time 02-06-18 2:50 PM
Received by: JP
Relinquished by: CP

Date/Time 02-06-18 4:00 PM
Received by: CP
Relinquished by: JP