



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93124

Temp.: 23.5 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/07/2018	17:40
2	CAL2	1	Water			7.01	02/06/2018	17:41
3	CAL3	1	Water			10.03	02/06/2018	17:42
4	ICV	1	Water			7.00	02/06/2018	17:43
5	CCV1	1	Water			2.01	02/07/2018	17:45
6	J1400-02	1	Solid	20.02	20	8.91	02/06/2018	17:46
7	J1400-02DUP	1	Solid	20.03	20	8.92	02/06/2018	17:48
8	J1400-04	1	Solid	20.04	20	9.56	02/06/2018	17:50
9	J1400-06	1	Solid	20.02	20	8.90	02/06/2018	17:51
10	J1403-02	1	Solid	20.03	20	10.82	02/06/2018	18:00
11	J1461-01	1	Solid	20.02	20	8.63	02/06/2018	18:01
12	J1461-02	1	Solid	20.04	20	8.42	02/06/2018	18:02
13	J1461-03	1	Solid	20.00	20	8.97	02/06/2018	18:03
14	J1461-04	1	Solid	20.03	20	6.81	02/06/2018	18:05
15	J1465-03	1	Solid	20.02	20	7.61	02/06/2018	18:06
16	CCV2	1	Water			12.02	02/06/2018	18:09

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY J1464

WorkList ID : 108289

Date : 2/6/2018 2:52:46 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/06/2018	Solid	J1400-02	Corrosivity	Cool 4 deg C	PSEG05	J63	EO-01-020618-A	02/06/2018	9045C
02/06/2018	Solid	J1400-04	Corrosivity	Cool 4 deg C	PSEG05	J63	EO-01-020618-B	02/06/2018	9045C
02/06/2018	Solid	J1400-06	Corrosivity	Cool 4 deg C	PSEG05	J63	EO-01-020618-C	02/06/2018	9045C
02/06/2018	Solid	J1403-02	Corrosivity	Cool 4 deg C	PSEG05	J63	SU-04-020618	02/06/2018	9045C
02/05/2018	Solid	J1461-01	Corrosivity	Cool 4 deg C	CLEA11	O33	CHASSIS-WASH	02/05/2018	9045C
02/05/2018	Solid	J1461-02	Corrosivity	Cool 4 deg C	CLEA11	O33	BUS-WASH	02/05/2018	9045C
02/05/2018	Solid	J1461-03	Corrosivity	Cool 4 deg C	CLEA11	O33	OIL-WATER-SEPERATOR	02/05/2018	9045C
02/05/2018	Solid	J1461-04	Corrosivity	Cool 4 deg C	CLEA11	O33	RECLAIMER-SLUDGE-PIT	02/05/2018	9045C
02/06/2018	Solid	J1465-03	Corrosivity	Cool 4 deg C	PSEG01	J63	RT-3451	02/06/2018	9045C

Date/Time

Received by:

Relinquished by:

Date/Time

Received by:

Relinquished by: