

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 02/07/2018

DATE & TIME IN: 02/07/2018 08:30

OVEN TEMP: 103 °C

Run Number: LB93127

DATE & TIME OUT: 02/07/2018 13:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB93127BL	LB93127BL	1.8978	1.8978	100	1.8978	1.8978	1.8978	0	0
2	LB93127BS	LB93127BS	1.8857	1.8857	100	1.9395	1.9395	1.9395	0.0538	538
3	J1466-01	FRAC-TANK	1.8657	1.8657	800	1.8721	1.8721	1.8721	0.0064	8

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS020718

WorkList ID : 108303

Date : 2/7/2018 7:59:07 AM

2893127

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/09/2018	Water	J1466-01	TSS	Cool 4 deg C	PSEG01	J63	FRAC-TANK	02/06/2018	SM2540 D

Date/Time 02/07/18, 08:05
 Received by: AS
 Relinquished by: AS

Date/Time 02/07/18, 08:35
 Received by: AS
 Relinquished by: AS