


7471B

LB 93128
INSTRUMENT ID: CY1

Linear

 μ Abs.:

41676



A= 0.0000e+000

B= 2.4278e-004

C= 1.9393e-002

Rho= 0.9996582

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.020	0.020	1	0.000	1				
0.2	0.200	0.214	0.014	802	0.0 %	802				
2.5	2.500	2.561	0.061	10467	0.0 %	10467				
5.0	5.000	4.939	-0.061	20262	0.0 %	20262				
7.5	7.500	7.330	-0.170	30112	0.0 %	30112				
10	10.000	10.137	0.137	41676	0.0 %	41676				

LB93128 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Feb 2018 10:05:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	1	-	0.0000	7471B	PPB	07 Feb 2018 10:20:36	S	Std
0.2 - 1	S0.2	802	-	0.2000	7471B	PPB	07 Feb 2018 10:22:53	S	Std
2.5 - 1	S2.5	10467	-	2.5000	7471B	PPB	07 Feb 2018 10:25:11	S	Std
5.0 - 1	S5	20262	-	5.0000	7471B	PPB	07 Feb 2018 10:27:29	S	Std
7.5 - 1	S7.5	30112	-	7.5000	7471B	PPB	07 Feb 2018 10:29:46	S	Std
10 - 1	S10	41676	-	10.0000	7471B	PPB	07 Feb 2018 10:32:02	S	Std
ICV87 - 1	ICV87	16085	3.9244	-	7471B	PPB	07 Feb 2018 10:35:20	U	SMPL
ICB87 - 1	ICB87	-103	-0.0056	-	7471B	PPB	07 Feb 2018 10:37:35	U	SMPL
CCV85 - 1	CCV85	20344	4.9584	-	7471B	PPB	07 Feb 2018 10:39:51	U	SMPL
CCB85 - 1	CCB85	-110	-0.0073	-	7471B	PPB	07 Feb 2018 10:42:07	U	SMPL
CRA - 1	CRA	800	0.2136	-	7471B	PPB	07 Feb 2018 10:44:23	U	SMPL
HighStd - 1	HighStd	41768	10.1597	-	7471B	PPB	07 Feb 2018 10:46:39	U	SMPL
ChkStd - 1	ChkStd	29090	7.0818	-	7471B	PPB	07 Feb 2018 10:48:55	U	SMPL
PB106345BL - 1	PBS	-126	-0.0112	-	7471B	PPB	07 Feb 2018 10:51:13	U	SMPL
PB106345BS - 1	LCSS	15944	3.8902	-	7471B	PPB	07 Feb 2018 10:53:31	U	SMPL
J1400-01 - 1	EO-01-020618-A	4059	1.0048	-	7471B	PPB	07 Feb 2018 10:55:48	U	SMPL
J1400-03 - 1	EO-01-020618-B	28007	6.8188	-	7471B	PPB	07 Feb 2018 10:58:06	U	SMPL
J1400-05 - 1	EO-01-020618-C	827	0.2202	-	7471B	PPB	07 Feb 2018 11:00:23	U	SMPL
J1403-01 - 1	SU-04-020618	1493	0.3819	-	7471B	PPB	07 Feb 2018 11:02:42	U	SMPL
J1465-01 - 1	RT-3451	743	0.1998	-	7471B	PPB	07 Feb 2018 11:04:58	U	SMPL
CCV86 - 1	CCV86	20414	4.9754	-	7471B	PPB	07 Feb 2018 11:07:14	U	SMPL
CCB86 - 1	CCB86	-96	-0.0039	-	7471B	PPB	07 Feb 2018 11:09:30	U	SMPL
J1465-01DUP - 1	RT-3451DUP	618	0.1694	-	7471B	PPB	07 Feb 2018 11:11:47	U	SMPL
J1465-01MS - 1	RT-3451MS	17188	4.1922	-	7471B	PPB	07 Feb 2018 11:14:03	U	SMPL
J1465-01MSD - 1	RT-3451MSD	17019	4.1512	-	7471B	PPB	07 Feb 2018 11:16:19	U	SMPL
J1468-01 - 1	EP-27	1318	0.3394	-	7471B	PPB	07 Feb 2018 11:18:35	U	SMPL
J1468-02 - 1	EP-30	3011	0.7504	-	7471B	PPB	07 Feb 2018 11:20:52	U	SMPL
J1468-03 - 1	EP-33	340	0.1019	-	7471B	PPB	07 Feb 2018 11:23:08	U	SMPL
J1468-04 - 1	EP-31	1097	0.2857	-	7471B	PPB	07 Feb 2018 11:25:26	U	SMPL
J1468-05 - 1	EP-28	2341	0.5877	-	7471B	PPB	07 Feb 2018 11:27:44	U	SMPL
J1468-06 - 1	EP-25	504	0.1418	-	7471B	PPB	07 Feb 2018 11:30:02	U	SMPL
J1468-07 - 1	EP-34	122	0.0490	-	7471B	PPB	07 Feb 2018 11:32:18	U	SMPL
CCV87 - 1	CCV87	20242	4.9337	-	7471B	PPB	07 Feb 2018 11:34:34	U	SMPL
CCB87 - 1	CCB87	-89	-0.0022	-	7471B	PPB	07 Feb 2018 11:36:51	U	SMPL
J1468-08 - 1	EP-B3	3354	0.8337	-	7471B	PPB	07 Feb 2018 11:39:07	U	SMPL
J1468-09 - 1	EP-35	784	0.2097	-	7471B	PPB	07 Feb 2018 11:41:23	U	SMPL
J1465-01LX5 - 1		131	0.0512	-	7471B	PPB	07 Feb 2018 11:43:39	U	SMPL
J1465-01A - 1		17307	4.2211	-	7471B	PPB	07 Feb 2018 11:45:56	U	SMPL
CCV88 - 1	CCV88	20546	5.0075	-	7471B	PPB	07 Feb 2018 11:50:45	U	SMPL
CCB88 - 1	CCB88	-121	-0.0100	-	7471B	PPB	07 Feb 2018 11:53:01	U	SMPL