



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93129

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 99.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/07/2018	09:00
2	CAL2	1	Water			7.00	02/07/2018	09:02
3	CAL3	1	Water			10.01	02/07/2018	09:04
4	ICV	1	Water			7.00	02/07/2018	10:40
5	CCV1	1	Water			2.00	02/07/2018	10:42
6	J1473-04	1	Solid	20	20	7.97	02/07/2018	10:44
7	J1473-04DUP	1	Solid	20.01	20	7.98	02/07/2018	10:46
8	J1473-08	1	Solid	20.03	20	7.71	02/07/2018	10:48
9	J1473-12	1	Solid	20.05	20	8.03	02/07/2018	10:50
10	J1473-16	1	Solid	20	20	8.12	02/07/2018	10:52
11	J1473-20	1	Solid	20.01	20	8.21	02/07/2018	10:54
12	J1473-24	1	Solid	20.02	20	8.26	02/07/2018	10:56
13	J1473-28	1	Solid	20	20	7.91	02/07/2018	10:58
14	J1473-32	1	Solid	20.01	20	7.87	02/07/2018	11:00
15	J1473-36	1	Solid	20.03	20	9.22	02/07/2018	11:02
16	CCV2	1	Water			12.00	02/07/2018	11:04
17	J1473-40	1	Solid	20.05	20	7.94	02/07/2018	11:06
18	J1473-44	1	Solid	20	20	7.46	02/07/2018	11:08
19	CCV3	1	Water			2.00	02/07/2018	11:10

WORKLIST(Hardcopy Internal Chain)

LB93129

WorkList Name : corrosivity020717

WorkList ID : 108327

Date : 2/7/2018 10:25:43 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/06/2018	Solid	J1473-04	Corrosivity	Cool 4 deg C	PSEG01	J63	B1	02/06/2018	9045C
02/06/2018	Solid	J1473-08	Corrosivity	Cool 4 deg C	PSEG01	J63	B2	02/06/2018	9045C
02/06/2018	Solid	J1473-12	Corrosivity	Cool 4 deg C	PSEG01	J63	B3	02/06/2018	9045C
02/06/2018	Solid	J1473-16	Corrosivity	Cool 4 deg C	PSEG01	J63	B4	02/06/2018	9045C
02/06/2018	Solid	J1473-20	Corrosivity	Cool 4 deg C	PSEG01	J63	B5	02/06/2018	9045C
02/06/2018	Solid	J1473-24	Corrosivity	Cool 4 deg C	PSEG01	J63	B6	02/06/2018	9045C
02/06/2018	Solid	J1473-28	Corrosivity	Cool 4 deg C	PSEG01	J63	B7	02/06/2018	9045C
02/06/2018	Solid	J1473-32	Corrosivity	Cool 4 deg C	PSEG01	J63	B8	02/06/2018	9045C
02/06/2018	Solid	J1473-36	Corrosivity	Cool 4 deg C	PSEG01	J63	B9	02/06/2018	9045C
02/06/2018	Solid	J1473-40	Corrosivity	Cool 4 deg C	PSEG01	J63	B10	02/06/2018	9045C
02/06/2018	Solid	J1473-44	Corrosivity	Cool 4 deg C	PSEG01	J63	B11	02/06/2018	9045C

Date/Time 02/07/18 10:35 AM

Received by: SP

Relinquished by: SP

Date/Time 02/07/18 11:20 AM

Received by: SP

Relinquished by: SP