

Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB93132

Reagent/Standard	Lot/Log #
MBAS CAL STD, 0.5PPM	WP61678
MBAS CCV STD, 1.0PPM	WP61683
MBAS DAILY CHECK STD, 0.01PPM	WP61685
MBAS DAILY CHECK STD, 2.0PPM	WP61686
MBAS CAL STD 2 PPM	WP61676
MBAS CAL STD 1 PPM	WP61675
MBAS CAL STD 0.1 PPM	WP61681
mbas cal std 0.05 ppm	WP61682
MBAS CAL STD 0.01 PPM	WP61689
MBAS BLANK 0 PPM	WP61674
MBAS CAL STD 0.4 PPM	WP61679
MBAS CAL STD 0.2 PPM	WP61680
MBAS CCV STD, 1.0PPM	WP61683
MBAS ICV-LCS std 1.0PPM	WP61684

Intercept: -0.0013

Slope: 0.2271

Regression: 0.9998

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Anal Date	Anal Time
1	CAL1	0	1	0.000	02/07/2018	11:12
2	CAL2	0.01	1	0.002	02/07/2018	11:12
3	CAL3	0.05	1	0.011	02/07/2018	11:13
4	CAL4	0.1	1	0.024	02/07/2018	11:13
5	CAL5	0.2	1	0.045	02/07/2018	11:14
6	CAL6	0.4	1	0.088	02/07/2018	11:15
7	CAL7	0.5	1	0.106	02/07/2018	11:16
8	CAL8	1	1	0.221	02/07/2018	11:16
9	CAL9	2	1	0.455	02/07/2018	11:17



Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB93132

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Result (mg/L)	AnalDate	Anal Time
1	ICV	1	1	0.216	0.957	02/07/2018	11:17
2	ICB		1	0.000	0.006	02/07/2018	11:18
3	CHK1	0.01	1	0.003	0.019	02/07/2018	11:18
4	CHK2	2	1	0.451	1.992	02/07/2018	11:19
5	CCV1	1	1	0.220	0.974	02/07/2018	11:19
6	CCB1		1	0.000	0.006	02/07/2018	11:20
7	LB93132BL		1	0.000	0.006	02/07/2018	11:20
8	LB93132BS	1	1	0.215	0.952	02/07/2018	11:21
9	J1161-07		1	0.012	0.059	02/07/2018	11:21
10	J1161-08		1	0.007	0.037	02/07/2018	11:22
11	J1161-09		1	0.011	0.054	02/07/2018	11:22
12	J1192-26		1	0.127	0.565	02/07/2018	11:23
13	CCV2	1	1	0.219	0.970	02/07/2018	11:23
14	CCB2		1	0.000	0.006	02/07/2018	11:24