

7470A


L393134

INSTRUMENT ID: CVI

Linear

 μ Abs.:

36422



A= 0.0000e+000

B= 2.7473e-004

C= -4.5872e-004

Rho= 0.9999868

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.025	0.025	93	0.000	93				
0.2	0.200	0.202	0.002	736	0.0 %	736				
2.5	2.500	2.462	-0.038	8962	0.0 %	8962				
5.0	5.000	5.003	0.003	18212	0.0 %	18212				
7.5	7.500	7.503	0.003	27313	0.0 %	27313				
10.0	10.000	10.006	0.006	36422	0.0 %	36422				

LB93134 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Feb 2018 14:04:56

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	93	-	0.0000	7470A	PPB	07 Feb 2018 14:09:47	S	Std
0.2 - 1	50.2	736	-	0.2000	7470A	PPB	07 Feb 2018 14:11:55	S	Std
2.5 - 1	52.5	8962	-	2.5000	7470A	PPB	07 Feb 2018 14:14:03	S	Std
5.0 - 1	55	18212	-	5.0000	7470A	PPB	07 Feb 2018 14:16:12	S	Std
7.5 - 1	57.5	27313	-	7.5000	7470A	PPB	07 Feb 2018 14:18:20	S	Std
10.0 - 1	510	36422	-	10.0000	7470A	PPB	07 Feb 2018 14:20:27	S	Std
ICV88 - 1	ICV88	14141	3.8844	-	7470A	PPB	07 Feb 2018 14:23:02	U	SMPL
ICB88 - 1	ICB88	-54	-0.0153	-	7470A	PPB	07 Feb 2018 14:25:08	U	SMPL
CCV89 - 1	CCV89	18327	5.0344	-	7470A	PPB	07 Feb 2018 14:27:15	U	SMPL
CCB89 - 1	CCB89	-35	-0.0101	-	7470A	PPB	07 Feb 2018 14:29:21	U	SMPL
CRA - 1	CRA	685	0.1877	-	7470A	PPB	07 Feb 2018 14:31:28	U	SMPL
HighStd - 1	HighStd	35836	9.8446	-	7470A	PPB	07 Feb 2018 14:33:35	U	SMPL
ChkStd - 1	ChkStd	25118	6.9001	-	7470A	PPB	07 Feb 2018 14:35:42	U	SMPL
PB106344BL - 1	PBW	-103	-0.0288	-	7470A	PPB	07 Feb 2018 14:37:51	U	SMPL
PB106344BS - 1	LCSW	12638	3.4715	-	7470A	PPB	07 Feb 2018 14:40:00	U	SMPL
PB106344TB - 1	PB106316TB	-26	-0.0076	-	7470A	PPB	07 Feb 2018 14:42:07	U	SMPL
J1459-04 - 1	TP-A	53	0.0141	-	7470A	PPB	07 Feb 2018 14:44:16	U	SMPL
J1459-04DUP - 1	TP-ADUP	68	0.0182	-	7470A	PPB	07 Feb 2018 14:46:24	U	SMPL
J1459-04MS - 1	TP-AMS	13350	3.6671	-	7470A	PPB	07 Feb 2018 14:48:33	U	SMPL
J1459-04MSD - 1	TP-AMSD	13476	3.7017	-	7470A	PPB	07 Feb 2018 14:50:39	U	SMPL
CCV90 - 1	CCV90	18538	5.0924	-	7470A	PPB	07 Feb 2018 14:52:46	U	SMPL
CCB90 - 1	CCB90	-44	-0.0125	-	7470A	PPB	07 Feb 2018 14:54:53	U	SMPL
J1459-08 - 1	TP-B	81	0.0218	-	7470A	PPB	07 Feb 2018 14:57:00	U	SMPL
J1460-04 - 1	B1	43	0.0114	-	7470A	PPB	07 Feb 2018 14:59:07	U	SMPL
J1460-08 - 1	B2	45	0.0119	-	7470A	PPB	07 Feb 2018 15:01:14	U	SMPL
PB106346BL - 1	PBW	40	0.0105	-	7470A	PPB	07 Feb 2018 15:03:21	U	SMPL
PB106346BS - 1	LCSW	13325	3.6603	-	7470A	PPB	07 Feb 2018 15:05:28	U	SMPL
J1466-01 - 1	FRAC-TANK	-229	-0.0634	-	7470A	PPB	07 Feb 2018 15:07:36	U	SMPL
J1466-01DUP - 1	FRAC-TANKDUP	-151	-0.0419	-	7470A	PPB	07 Feb 2018 15:09:44	U	SMPL
J1466-01MS - 1	FRAC-TANKMS	11930	3.2770	-	7470A	PPB	07 Feb 2018 15:11:53	U	SMPL
J1466-01MSD - 1	FRAC-TANKMSD	11973	3.2888	-	7470A	PPB	07 Feb 2018 15:14:01	U	SMPL
CCV91 - 1	CCV91	18227	5.0070	-	7470A	PPB	07 Feb 2018 15:16:08	U	SMPL
CCB91 - 1	CCB91	-24	-0.0071	-	7470A	PPB	07 Feb 2018 15:18:16	U	SMPL
J1459-04LX5 - 1		42	0.0111	-	7470A	PPB	07 Feb 2018 15:20:24	U	SMPL
J1459-04A - 1		13801	3.7910	-	7470A	PPB	07 Feb 2018 15:22:31	U	SMPL
J1466-01LX5 - 1		-74	-0.0208	-	7470A	PPB	07 Feb 2018 15:24:38	U	SMPL
J1466-01A - 1		12619	3.4663	-	7470A	PPB	07 Feb 2018 15:26:45	U	SMPL
CCV92 - 1	CCV92	18342	5.0386	-	7470A	PPB	07 Feb 2018 15:28:53	U	SMPL
CCB92 - 1	CCB92	-67	-0.0189	-	7470A	PPB	07 Feb 2018 15:31:00	U	SMPL