



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB93145

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	02/07/2018	08:45
2	J1466-02		1	>100.0	>212.0	02/07/2018	09:20
3	J1466-02DUP		1	>100.0	>212.0	02/07/2018	09:45

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP2718

WorkList ID : 108307

Date : 2/7/2018 8:02:12 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/09/2018	Water	J1466-02 B	Flash Point	Cool 4 deg C	PSEG01	J63	FRAC-TANK	02/06/2018	1010A

LB 93145

Date/Time 2/7/18 8:15AM
 Received by: KS
 Relinquished by: CP

Date/Time 2/7/18 09:50 AM
 Received by: CP
 Relinquished by: KS