

245.1


LB 93150

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37427



A= 0.0000e+000

B= 2.7136e-004

C= -2.1455e-002

Rho= 0.9996640

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	115	0.000	115				
0.05	0.050	0.044	-0.006	243	0.0 %	243				
0.2	0.200	0.215	0.015	870	0.0 %	870				
2.5	2.500	2.549	0.049	9474	0.0 %	9474				
5.0	5.000	5.006	0.006	18528	0.0 %	18528				
7.5	7.500	7.291	-0.209	26946	0.0 %	26946				
10.0	10.000	10.135	0.135	37427	0.0 %	37427				

LB93150 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 07 Feb 2018 16:32:03

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc/Method	Units	Date	Type	Type
0.0 - 1	50	115	-	0.0000245.1	PPB	07 Feb 2018 16:35:50	S	Std
0.05 - 1	50.05	243	-	0.0500245.1	PPB	07 Feb 2018 16:37:59	S	Std
0.2 - 1	50.2	870	-	0.2000245.1	PPB	07 Feb 2018 16:40:07	S	Std
2.5 - 1	52.5	9474	-	2.5000245.1	PPB	07 Feb 2018 16:42:15	S	Std
5.0 - 1	55	18528	-	5.0000245.1	PPB	07 Feb 2018 16:44:23	S	Std
7.5 - 1	57.5	26946	-	7.5000245.1	PPB	07 Feb 2018 16:51:44	S	Std
10.0 - 1	510	37427	-	10.0000245.1	PPB	07 Feb 2018 16:59:04	S	Std
ICV89 - 1	ICV89	14357	3.8745	-245.1	PPB	07 Feb 2018 17:01:53	U	SMPL
ICB89 - 1	ICB89	8	-0.0193	-245.1	PPB	07 Feb 2018 17:03:59	U	SMPL
CCV93 - 1	CCV93	18592	5.0237	-245.1	PPB	07 Feb 2018 17:06:05	U	SMPL
CCB93 - 1	CCB93	0	-0.0215	-245.1	PPB	07 Feb 2018 17:08:12	U	SMPL
CRA - 1	CRA	865	0.2133	-245.1	PPB	07 Feb 2018 17:10:19	U	SMPL
HighStd - 1	HighStd	37119	10.0512	-245.1	PPB	07 Feb 2018 17:12:25	U	SMPL
ChkStd - 1	ChkStd	24660	6.6703	-245.1	PPB	07 Feb 2018 17:14:32	U	SMPL
PB106353BL - 1	PBW	-69	-0.0402	-245.1	PPB	07 Feb 2018 17:16:41	U	SMPL
PB106353BS - 1	LCSW	12829	3.4598	-245.1	PPB	07 Feb 2018 17:18:49	U	SMPL
J1161-10 - 1	LOD-MDL-WATER-04-QT1-2018	269	0.0515	-245.1	PPB	07 Feb 2018 17:28:57	U	SMPL
J1161-11 - 1	LOQ-WATER-05-QT1-2018	1417	0.3631	-245.1	PPB	07 Feb 2018 17:31:05	U	SMPL
J1161-12 - 1	MDL-WATER-06-QT1-2018	250	0.0464	-245.1	PPB	07 Feb 2018 17:33:13	U	SMPL
J1455-01 - 1	LW-01-B	-59	-0.0375	-245.1	PPB	07 Feb 2018 17:35:22	U	SMPL
J1455-01DUP - 1	LW-01-BDUP	-61	-0.0380	-245.1	PPB	07 Feb 2018 17:37:28	U	SMPL
CCV94 - 1	CCV94	18423	4.9778	-245.1	PPB	07 Feb 2018 17:39:35	U	SMPL
CCB94 - 1	CCB94	-3	-0.0223	-245.1	PPB	07 Feb 2018 17:41:42	U	SMPL
J1455-02 - 1	J1455-01MS	13906	3.7521	-245.1	PPB	07 Feb 2018 17:43:48	U	SMPL
J1455-03 - 1	J1455-01MSD	13459	3.6308	-245.1	PPB	07 Feb 2018 17:45:55	U	SMPL
J1455-04 - 1	LW-01-OW	-164	-0.0660	-245.1	PPB	07 Feb 2018 17:48:02	U	SMPL
J1455-05 - 1	LW-02-B	-145	-0.0608	-245.1	PPB	07 Feb 2018 17:50:09	U	SMPL
J1455-06 - 1	LW-02-OW	-46	-0.0339	-245.1	PPB	07 Feb 2018 17:52:16	U	SMPL
J1455-07 - 1	LW-03-B	-84	-0.0442	-245.1	PPB	07 Feb 2018 17:54:24	U	SMPL
J1455-08 - 1	LW-03-OW	-77	-0.0423	-245.1	PPB	07 Feb 2018 17:56:32	U	SMPL
J1455-09 - 1	LW-04-B	-143	-0.0603	-245.1	PPB	07 Feb 2018 17:58:40	U	SMPL
J1455-01LX5 - 1		62	-0.0046	-245.1	PPB	07 Feb 2018 18:00:49	U	SMPL
J1455-01A - 1		14098	3.8042	-245.1	PPB	07 Feb 2018 18:02:56	U	SMPL
CCV95 - 1	CCV95	18396	4.9705	-245.1	PPB	07 Feb 2018 18:05:02	U	SMPL
CCB95 - 1	CCB95	-3	-0.0223	-245.1	PPB	07 Feb 2018 18:07:09	U	SMPL