



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93151

Temp.: 22.1 °C

BalanceID: WC SC-4

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	02/07/2018	16:30
2	CAL2	1	Water			7.01	02/07/2018	16:31
3	CAL3	1	Water			10.02	02/07/2018	16:32
4	ICV	1	Water			7.00	02/07/2018	16:34
5	CCV1	1	Water			2.01	02/07/2018	16:35
6	J1404-02	1	Solid	20.02	20	8.32	02/07/2018	16:36
7	J1404-02DUP	1	Solid	20.03	20	8.33	02/07/2018	16:37
8	J1485-02	1	Solid	20.05	20	8.58	02/07/2018	16:38
9	J1485-04	1	Solid	20.03	20	8.61	02/07/2018	16:40
10	CCV2	1	Water			12.01	02/07/2018	16:41

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 11487

WorkList ID : 108338

Date : 2/7/2018 4:17:17 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/07/2018	Solid	J1404-02	Corrosivity	Cool 4 deg C	PSEG05	M63	TR-05-020718	02/07/2018	9045C
02/07/2018	Solid	J1485-02	Corrosivity	Cool 4 deg C	PSEG01	M63	RO-201	02/07/2018	9045C
02/07/2018	Solid	J1485-04	Corrosivity	Cool 4 deg C	PSEG01	M63	RO-215	02/07/2018	9045C

Date/Time 02-07-18 4.20PM
 Received by: JP
 Relinquished by: JP

Date/Time 02-07-18 6.30PM
 Received by: JP
 Relinquished by: JP