



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93152

Temp.: 22.1 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	02/07/2018	16:30
2	CAL2	1	Water			7.01	02/07/2018	16:31
3	CAL3	1	Water			10.02	02/07/2018	16:32
4	ICV	1	Water			7.00	02/07/2018	16:35
5	CCV1	1	Water			2.02	02/07/2018	16:45
6	J1490-07	1	Water			7.77	02/07/2018	16:46
7	J1490-07DUP	1	Water			7.78	02/07/2018	16:47
8	CCV2	1	Water			12.02	02/07/2018	16:49

1503152

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY j1490 WorkList ID : 108339 Date : 2/7/2018 4:38:11 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/07/2018	Water	J1490-07	Corrosivity	Cool 4 deg C	PSEG02	O62	265578-SEGMENT-2C	02/07/2018	9040C

Date/Time 02-07-18 4:40 PM
Received by: JS
Relinquished by: JS

Date/Time 02-07-18 6:00 PM
Received by: JS
Relinquished by: JS