



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB93156
Analysis Date: 02/08/2018
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 02/08/2018
Extraction IN Time: 08:25
Extraction OUT Time: 09:15

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB93156BL	LB93156BL	WATER	2	1000	100	3.5497	3.5497	3.03	3.5497	3.5497	0	0
2	LB93156BS	LB93156BS	WATER	2	1000	100	3.0211	3.0211	3.02	3.0396	3.0396	0.0185	18.5
3	LB93156BSD	LB93156BSD	WATER	2	1000	100	3.9688	3.9688	3.01	3.9875	3.9875	0.0187	18.7
4	J1489-01	DSN-001-SEWER-OUTFALL	WATER	2	980	100	3.057	3.057	3.03	3.0575	3.0575	0.0005	0.51
5	J1489-02	DSN-002-SEWER-OUTFALL	WATER	2	970	100	3.0365	3.0365	3.04	3.0367	3.0367	0.0002	0.21



QC Batch# LB93156 Test: Non-Polar Ma Analysis Date: 02/08/2018

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2317
pH Paper 0-14	
Sodium Sulfate	EP1822
1:1 HCL	WP59249
Silica Gel	W2199
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP58006
LCSWD	10.00 ML	WP58005
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0019 In OVEN TEMP : 70 Dessicator Time In : 10:15
1.0000 gram Balance: 1.0003 In Time: 09:35
Bal Check Time: 08:45

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 71 Dessicator Time Out: 12:00
1.0000 gram Balance: 1.0005 Out Time: 11:10
Bal Check Time: 12:35

13 93156

WORKLIST(Hardcopy Internal Chain)

WorkList Name : NON POPLAR J1489 WorkList ID : 108359 Date : 2/8/2018 8:09:38 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/20/2018	Water	J1489-01 - <u>6</u>	Non-Polar Material	Conc H2SO4 tk	TRIS02	O52	DSN-001-SEWER-OUTFALL-1	02/07/2018	1664A
02/20/2018	Water	J1489-02 - <u>6</u>	Non-Polar Material	Conc H2SO4 tk	TRIS02	O52	DSN-002-SEWER-OUTFALL-1	02/07/2018	1664A

Date/Time 02-08-18 8:25 AM
 Received by: SP
 Relinquished by: SP

Date/Time 02-08-18 11:30 AM
 Received by: SP
 Relinquished by: SP